

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
 Data File : VN082887.D
 Acq On : 11 Jul 2024 19:35
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 12 01:11:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	74	0.00
2 T	Dichlorodifluoromethane	0.647	0.695	-7.4	76	0.00
3 P	Chloromethane	0.614	0.648	-5.5	82	0.00
4 C	Vinyl Chloride	0.645	0.708	-9.8#	81	0.00
5 T	Bromomethane	0.465	0.360	22.6	61	0.00
6 T	Chloroethane	0.441	0.497	-12.7	86	0.00
7 T	Trichlorofluoromethane	0.927	1.013	-9.3	85	0.00
8 T	Diethyl Ether	0.317	0.372	-17.4	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.525	0.572	-9.0	80	-0.01
10 T	Methyl Iodide	0.538	0.426	20.8	53	0.00
11 T	Tert butyl alcohol	0.118	0.129	-9.3	82	0.00
12 CM	1,1-Dichloroethene	0.510	0.545	-6.9#	79	0.00
13 T	Acrolein	0.085	0.088	-3.5	78	0.00
14 T	Allyl chloride	0.796	0.859	-7.9	85	0.00
15 T	Acrylonitrile	0.271	0.319	-17.7	89	0.00
16 T	Acetone	0.247	0.275	-11.3	86	0.00
17 T	Carbon Disulfide	1.525	1.510	1.0	77	0.00
18 T	Methyl Acetate	0.653	0.782	-19.8	94	0.00
19 T	Methyl tert-butyl Ether	1.761	2.050	-16.4	85	0.00
20 T	Methylene Chloride	0.596	0.661	-10.9	88	0.00
21 T	trans-1,2-Dichloroethene	0.553	0.586	-6.0	81	0.00
22 T	Diisopropyl ether	1.665	2.069	-24.3	92	0.00
23 T	Vinyl Acetate	1.322	1.657	-25.3#	90	0.00
24 P	1,1-Dichloroethane	1.010	1.184	-17.2	87	0.00
25 T	2-Butanone	0.373	0.434	-16.4	88	0.00
26 T	2,2-Dichloropropane	0.939	0.919	2.1	71	0.00
27 T	cis-1,2-Dichloroethene	0.647	0.725	-12.1	84	0.00
28 T	Bromochloromethane	0.433	0.526	-21.5	95	0.00
29 T	Tetrahydrofuran	0.235	0.287	-22.1	91	0.00
30 C	Chloroform	1.071	1.251	-16.8#	88	0.00
31 T	Cyclohexane	0.916	0.858	6.3	76	0.00
32 T	1,1,1-Trichloroethane	0.995	1.135	-14.1	86	0.00
33 S	1,2-Dichloroethane-d4	0.701	0.761	-8.6	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.321	0.336	-4.7	82	0.00
36 T	1,1-Dichloropropene	0.442	0.473	-7.0	81	0.00
37 T	Ethyl Acetate	0.450	0.517	-14.9	90	0.00
38 T	Carbon Tetrachloride	0.533	0.559	-4.9	80	0.00
39 T	Methylcyclohexane	0.478	0.459	4.0	72	0.00
40 TM	Benzene	1.361	1.507	-10.7	86	0.00
41 T	Methacrylonitrile	0.243	0.276	-13.6	89	0.00
42 TM	1,2-Dichloroethane	0.490	0.570	-16.3	90	0.00
43 T	Isopropyl Acetate	1.386	0.945	31.8#	87	0.00
44 TM	Trichloroethene	0.349	0.358	-2.6	80	0.00
45 C	1,2-Dichloropropane	0.318	0.372	-17.0#	89	0.00
46 T	Dibromomethane	0.247	0.280	-13.4	87	0.00
47 T	Bromodichloromethane	0.511	0.582	-13.9	89	0.00
48 T	Methyl methacrylate	0.353	0.418	-18.4	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
 Data File : VN082887.D
 Acq On : 11 Jul 2024 19:35
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 12 01:11:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.009	-28.6#	89	0.00
50 S	Toluene-d8	1.185	1.224	-3.3	79	0.00
51 T	4-Methyl-2-Pentanone	0.446	0.540	-21.1	92	0.00
52 CM	Toluene	0.850	0.954	-12.2#	84	0.00
53 T	t-1,3-Dichloropropene	0.527	0.589	-11.8	83	0.00
54 T	cis-1,3-Dichloropropene	0.558	0.617	-10.6	85	0.00
55 T	1,1,2-Trichloroethane	0.313	0.374	-19.5	90	0.00
56 T	Ethyl methacrylate	0.486	0.621	-27.8#	90	0.00
57 T	1,3-Dichloropropane	0.553	0.647	-17.0	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.336	-34.4#	97	0.00
59 T	2-Hexanone	0.332	0.413	-24.4	91	0.00
60 T	Dibromochloromethane	0.398	0.464	-16.6	88	0.00
61 T	1,2-Dibromoethane	0.332	0.384	-15.7	88	0.00
62 S	4-Bromofluorobenzene	0.426	0.439	-3.1	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.373	0.372	0.3	82	0.00
65 PM	Chlorobenzene	1.071	1.126	-5.1	84	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.417	-9.4	87	0.00
67 C	Ethyl Benzene	1.801	1.912	-6.2#	82	0.00
68 T	m/p-Xylenes	0.680	0.741	-9.0	82	0.00
69 T	o-Xylene	0.651	0.729	-12.0	85	0.00
70 T	Styrene	1.076	1.264	-17.5	85	0.00
71 P	Bromoform	0.300	0.339	-13.0	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
73 T	Isopropylbenzene	3.633	3.829	-5.4	81	0.00
74 T	N-amyl acetate	1.503	1.852	-23.2	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.088	1.211	-11.3	91	0.00
76 T	1,2,3-Trichloropropane	0.862	1.129	-31.0#	105	0.00
77 T	Bromobenzene	0.937	0.994	-6.1	85	0.00
78 T	n-propylbenzene	4.177	4.409	-5.6	80	0.00
79 T	2-Chlorotoluene	2.662	2.805	-5.4	84	0.00
80 T	1,3,5-Trimethylbenzene	2.937	3.154	-7.4	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.438	0.457	-4.3	75	0.00
82 T	4-Chlorotoluene	2.635	2.782	-5.6	84	0.00
83 T	tert-Butylbenzene	2.604	2.679	-2.9	79	0.00
84 T	1,2,4-Trimethylbenzene	2.993	3.240	-8.3	81	0.00
85 T	sec-Butylbenzene	3.518	3.599	-2.3	77	0.00
86 T	p-Isopropyltoluene	2.935	3.102	-5.7	78	0.00
87 T	1,3-Dichlorobenzene	1.693	1.756	-3.7	83	0.00
88 T	1,4-Dichlorobenzene	1.751	1.748	0.2	82	0.00
89 T	n-Butylbenzene	2.515	2.509	0.2	76	0.00
90 T	Hexachloroethane	0.621	0.617	0.6	79	0.00
91 T	1,2-Dichlorobenzene	1.644	1.687	-2.6	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.252	0.273	-8.3	89	0.00
93 T	1,2,4-Trichlorobenzene	0.886	0.877	1.0	76	0.00
94 T	Hexachlorobutadiene	0.393	0.344	12.5	71	0.00
95 T	Naphthalene	3.033	3.185	-5.0	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
Data File : VN082887.D
Acq On : 11 Jul 2024 19:35
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jul 12 01:11:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.911	0.893	2.0	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6