


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071218\
Data File : VN049767.D
Acq On : 12 Jul 2018 9:37
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
Title : SW846 8260
Last Update : Thu Jul 12 15:51:38 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3347

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.1	15391	PASS
75	95	30	60	48.1	41026	PASS
95	95	100	100	100.0	85216	PASS
96	95	5	9	6.8	5807	PASS
173	174	0.00	2	0.6	422	PASS
174	95	50	100	89.1	75885	PASS
175	174	5	9	8.3	6290	PASS
176	174	95	101	96.6	73336	PASS
177	176	5	9	7.8	5733	PASS