

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071219\
 Data File : VN056673.D
 Acq On : 11 Jul 2019 18:28
 Operator : JC/SP
 Sample : K3740-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-22

Quant Time: Jul 12 09:05:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	267020	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	466068	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	483982	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	176050	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	156909	53.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.04%	
35) Dibromofluoromethane	7.59	113	139577	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
50) Toluene-d8	10.09	98	588329	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
62) 4-Bromofluorobenzene	12.40	95	206924	54.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.74%	
Target Compounds						
16) Acetone	3.82	43	21452	14.468	ug/l	92
20) Methylene Chloride	4.55	84	144864	51.861	ug/l	99
30) Chloroform	7.38	83	5046	1.080	ug/l	85
43) Isopropyl Acetate	8.17	43	11086	1.783	ug/l #	90
95) Naphthalene	15.14	128	3741	5.660	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071219\
Data File : VN056673.D
Acq On : 11 Jul 2019 18:28
Operator : JC/SP
Sample : K3740-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-22

Quant Time: Jul 12 09:05:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

