


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN071219\
 Data File : VN056665.D
 Acq On : 11 Jul 2019 14:32
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Title : SW846 8260
 Last Update : Fri Jun 21 03:45:40 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	7063	PASS
75	95	30	60	47.2	19314	PASS
95	95	100	100	100.0	40877	PASS
96	95	5	9	6.5	2669	PASS
173	174	0.00	2	1.1	419	PASS
174	95	50	100	92.1	37648	PASS
175	174	5	9	7.0	2653	PASS
176	174	95	101	97.2	36605	PASS
177	176	5	9	6.4	2351	PASS