

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071223\
 Data File : VN078477.D
 Acq On : 12 Jul 2023 13:04
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 13 03:43:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	418481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	753976	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	688237	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	229823	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	281824	50.234	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.460%
35) Dibromofluoromethane	8.171	113	277596	54.237	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.480%
50) Toluene-d8	10.570	98	868873	44.938	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.880%
62) 4-Bromofluorobenzene	12.853	95	304288	51.093	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.180%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071223\
Data File : VN078477.D
Acq On : 12 Jul 2023 13:04
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 13 03:43:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1067
Delta R.T. 0.005 min
Lab File: VN078477.D
Acq: 12 Jul 2023 13:04




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 418481
Ion Ratio Lower Upper
168 100
99 58.6 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 50.234 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. 0.005 min
Lab File: VN078477.D
Acq: 12 Jul 2023 13:04

Tgt Ion: 65 Resp: 281824
Ion Ratio Lower Upper
65 100
67 53.0 0.0 106.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.112 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN078477.D

Acq: 12 Jul 2023 13:04

Instrument :

MSVOA_N

ClientSampleId :

IBLK





Tgt Ion:114 Resp: 753976

Ion Ratio Lower Upper

114 100

63 19.2 0.0 36.4

88 14.6 0.0 30.6



#35

Dibromofluoromethane

Concen: 54.237 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.006 min

Lab File: VN078477.D

Acq: 12 Jul 2023 13:04




Tgt Ion:113 Resp: 277596


Ion Ratio Lower Upper

113 100

111 104.0 82.6 123.8

192 17.5 12.7 19.1









#50
Toluene-d8
Concen: 44.938 ug/l
RT: 10.570 min Scan# 1465
Delta R.T. -0.001 min
Lab File: VN078477.D
Acq: 12 Jul 2023 13:04




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 868873
Ion Ratio Lower Upper
98 100
100 63.3 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 51.093 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VN078477.D
Acq: 12 Jul 2023 13:04

Tgt Ion: 95 Resp: 304288
Ion Ratio Lower Upper
95 100
174 77.1 0.0 143.8
176 74.1 0.0 135.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.870 min Scan# 1

Delta R.T. -0.001 min

Lab File: VN078477.D

Acq: 12 Jul 2023 13:04

Instrument :

MSVOA_N

ClientSampleId :

IBLK





Tgt Ion:117 Resp: 688237

Ion Ratio Lower Upper

117 100

82 53.8 44.0 66.0

119 32.3 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.800 min Scan# 2014

Delta R.T. -0.000 min

Lab File: VN078477.D

Acq: 12 Jul 2023 13:04

Tgt Ion:152 Resp: 229823

Ion Ratio Lower Upper

152 100

115 62.0 32.3 96.8

150 161.3 0.0 355.8

