

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071224\
 Data File : VN082894.D
 Acq On : 12 Jul 2024 13:03
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 15 01:10:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	102083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	194906	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	188773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	77589	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	79995	55.902	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.800%
35) Dibromofluoromethane	8.165	113	64018	51.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.380%
50) Toluene-d8	10.571	98	229463	49.675	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.340%
62) 4-Bromofluorobenzene	12.853	95	75829	45.700	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071224\
Data File : VN082894.D
Acq On : 12 Jul 2024 13:03
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 15 01:10:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

