

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071224\
 Data File : VN082896.D
 Acq On : 12 Jul 2024 13:51
 Operator : JC\MD
 Sample : P3229-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-189-TB-20240709

Quant Time: Jul 15 01:11:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	104333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	198275	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	178191	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	77326	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	81650	55.828	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.660%
35) Dibromofluoromethane	8.165	113	66479	52.252	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.500%
50) Toluene-d8	10.571	98	237154	50.468	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.940%
62) 4-Bromofluorobenzene	12.853	95	76356	45.235	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	90.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071224\
Data File : VN082896.D
Acq On : 12 Jul 2024 13:51
Operator : JC\MD
Sample : P3229-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-189-TB-20240709

Quant Time: Jul 15 01:11:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

