

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071224\
 Data File : VN082899.D
 Acq On : 12 Jul 2024 15:04
 Operator : JC\MD
 Sample : P3229-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-189-GW-800-802

Quant Time: Jul 15 01:12:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	88833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	171262	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	183457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	75912	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	69843	56.087	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.180%	
35) Dibromofluoromethane	8.171	113	58396	53.138	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.280%	
50) Toluene-d8	10.571	98	212293	52.303	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.600%	
62) 4-Bromofluorobenzene	12.847	95	67675	46.416	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 92.840%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071224\
Data File : VN082899.D
Acq On : 12 Jul 2024 15:04
Operator : JC\MD
Sample : P3229-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-189-GW-800-802

Quant Time: Jul 15 01:12:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

