

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071318\
 Data File : VN049801.D
 Acq On : 13 Jul 2018 17:01
 Operator : MD\SY
 Sample : PB111069ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111069ZHE#04

Quant Time: Jul 14 00:17:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 12 15:51:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1172122	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1868099	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1594336	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	611748	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	795936	58.37	ug/l	0.00
Spiked Amount			Recovery	=	116.74%	
35) Dibromofluoromethane	7.59	113	719866	51.12	ug/l	0.00
Spiked Amount			Recovery	=	102.24%	
50) Toluene-d8	10.09	98	2540711	52.71	ug/l	0.00
Spiked Amount			Recovery	=	105.42%	
62) 4-Bromofluorobenzene	12.40	95	724525	42.09	ug/l	0.00
Spiked Amount			Recovery	=	84.18%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	26617	5.90	ug/l	91
18) Methyl Acetate	4.34	43	62599	4.83	ug/l	98
20) Methylene Chloride	4.55	84	33636	1.52	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071318\
Data File : VN049801.D
Acq On : 13 Jul 2018 17:01
Operator : MD\SY
Sample : PB111069ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111069ZHE#04

Quant Time: Jul 14 00:17:20 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 12 15:51:38 2018
Response via : Initial Calibration

