

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071318\
 Data File : VN049810.D
 Acq On : 13 Jul 2018 20:55
 Operator : MD\SY
 Sample : PB111069ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111069ZHE#13

Quant Time: Jul 14 00:43:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 12 15:51:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1106591	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1839297	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1497168	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	566131	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	759465	59.00	ug/l	0.00
Spiked Amount			Recovery	=	118.00%	
35) Dibromofluoromethane	7.60	113	683343	49.29	ug/l	0.00
Spiked Amount			Recovery	=	98.58%	
50) Toluene-d8	10.09	98	2441997	51.46	ug/l	0.00
Spiked Amount			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	12.40	95	673935	39.76	ug/l	0.00
Spiked Amount			Recovery	=	79.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	28046	6.83	ug/l	99
18) Methyl Acetate	4.33	43	80025	6.81	ug/l	97
20) Methylene Chloride	4.56	84	29809	1.37	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071318\
Data File : VN049810.D
Acq On : 13 Jul 2018 20:55
Operator : MD\SY
Sample : PB111069ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111069ZHE#13

Quant Time: Jul 14 00:43:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 12 15:51:38 2018
Response via : Initial Calibration

