

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071318\
 Data File : VN049811.D
 Acq On : 13 Jul 2018 21:20
 Operator : MD\SY
 Sample : PB111069ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111069ZHE#01

Quant Time: Jul 14 00:47:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 12 15:51:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1165543	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1925112	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1588752	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	588122	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	807022	59.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.04%	
35) Dibromofluoromethane	7.60	113	724724	49.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.88%	
50) Toluene-d8	10.09	98	2597303	52.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.58%	
62) 4-Bromofluorobenzene	12.40	95	707666	39.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.78%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	25827	5.71	ug/l	89
18) Methyl Acetate	4.34	43	82107	6.61	ug/l	99
20) Methylene Chloride	4.56	84	30661	1.31	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071318\
Data File : VN049811.D
Acq On : 13 Jul 2018 21:20
Operator : MD\SY
Sample : PB111069ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111069ZHE#01

Quant Time: Jul 14 00:47:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 12 15:51:38 2018
Response via : Initial Calibration

