

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071318\
 Data File : VN049823.D
 Acq On : 14 Jul 2018 3:22
 Operator : MD\SY
 Sample : J3825-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-071118-D

Quant Time: Jul 14 04:33:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 12 15:51:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1177416	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1970705	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1584703	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	611325	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	811331	59.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.48%	
35) Dibromofluoromethane	7.59	113	742265	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
50) Toluene-d8	10.09	98	2603225	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
62) 4-Bromofluorobenzene	12.40	95	712372	39.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.46%	
Target Compounds						
16) Acetone	3.82	43	108926	30.49	ug/l	96
18) Methyl Acetate	4.34	43	78816	6.24	ug/l	98
20) Methylene Chloride	4.55	84	112937	7.25	ug/l	98
95) Naphthalene	15.14	128	197468	10.74	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071318\
Data File : VN049823.D
Acq On : 14 Jul 2018 3:22
Operator : MD\SY
Sample : J3825-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-02-071118-D

Quant Time: Jul 14 04:33:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 12 15:51:38 2018
Response via : Initial Calibration

