

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071318\
 Data File : VN049832.D
 Acq On : 14 Jul 2018 7:14
 Operator : MD\SY
 Sample : J3975-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-8

Manual Integrations
 APPROVED

MMDadoda
 7/16/2018 11:02:51 AM

Quant Time: Jul 16 00:57:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 12 15:51:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1218389	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1983746	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1631405	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	609708	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	826047	58.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.56%	
35) Dibromofluoromethane	7.59	113	752896	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
50) Toluene-d8	10.09	98	2649015	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
62) 4-Bromofluorobenzene	12.40	95	730833	39.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.96%	
Target Compounds						
16) Acetone	3.83	43	45861	11.16	ug/l #	88
18) Methyl Acetate	4.33	43	76709	5.83	ug/l	94
20) Methylene Chloride	4.52	84	45364414m	3177.25	ug/l	
25) 2-Butanone	6.85	43	11282	2.04	ug/l #	80
95) Naphthalene	15.14	128	78529	6.70	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071318\
Data File : VN049832.D
Acq On : 14 Jul 2018 7:14
Operator : MD\SY
Sample : J3975-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-8

Manual Integrations
APPROVED

MMDadoda
7/16/2018 11:02:51 AM

Quant Time: Jul 16 00:57:44 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071218W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 12 15:51:38 2018
Response via : Initial Calibration

