

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071319\
 Data File : VN056694.D
 Acq On : 12 Jul 2019 17:50
 Operator : JC/SP
 Sample : K3779-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 390711762-TRIP-BLANK

Quant Time: Jul 13 00:52:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N071319W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jul 12 23:58:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	49985	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	262494	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	212760	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	101060	29.85	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	99.50%
60) 4-Bromofluorobenzene	12.40	95	79963	24.56	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	81.87%
63) Toluene-d8	10.09	98	306346	31.51	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	105.03%

Target Compounds					Ovalue
15) Acetone	3.82	58	22658	37.759 ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071319\
Data File : VN056694.D
Acq On : 12 Jul 2019 17:50
Operator : JC/SP
Sample : K3779-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
390711762-TRIP-BLANK

Quant Time: Jul 13 00:52:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N071319W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jul 12 23:58:13 2019
Response via : Initial Calibration

