

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071321\
 Data File : VN067717.D
 Acq On : 13 Jul 2021 15:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jul 13 22:17:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	432214	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	771182	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	771438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	389621	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	313267	47.958	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.920%
35) Dibromofluoromethane	8.021	113	232050	49.804	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.600%
50) Toluene-d8	10.443	98	997862	51.165	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.340%
62) 4-Bromofluorobenzene	12.733	95	366601	53.212	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.420%

Target Compounds			Qvalue			
2) Dichlorodifluoromethane	2.070	85	169036	46.012	ug/l	100
3) Chloromethane	2.301	50	227693	39.132	ug/l	97
4) Vinyl Chloride	2.443	62	370636	42.206	ug/l	98
5) Bromomethane	2.832	94	307537	51.671	ug/l	96
6) Chloroethane	3.006	64	291607	48.944	ug/l	97
7) Trichlorofluoromethane	3.365	101	732080	55.173	ug/l	100
8) Diethyl Ether	3.819	74	129216	36.963	ug/l	91
9) 1,1,2-Trichlorotrifluo...	4.199	101	203361	50.769	ug/l	96
10) Methyl Iodide	4.414	142	246010	51.556	ug/l	93
11) Tert butyl alcohol	5.385	59	208002	221.595	ug/l	# 85
12) 1,1-Dichloroethene	4.170	96	187912	48.896	ug/l	90
13) Acrolein	4.036	56	125934	275.637	ug/l	99
14) Allyl chloride	4.835	41	334395	47.334	ug/l	93
15) Acrylonitrile	5.554	53	558490	239.885	ug/l	100
16) Acetone	4.285	43	453140	220.466	ug/l	95
17) Carbon Disulfide	4.521	76	498948	46.869	ug/l	98
18) Methyl Acetate	4.854	43	252560	43.748	ug/l	94
19) Methyl tert-butyl Ether	5.618	73	778399	48.877	ug/l	99
20) Methylene Chloride	5.093	84	237196	49.426	ug/l	91
21) trans-1,2-Dichloroethene	5.589	96	227690	50.111	ug/l	88
22) Diisopropyl ether	6.498	45	810560	50.562	ug/l	97
23) Vinyl Acetate	6.434	43	3405509	248.479	ug/l	96
24) 1,1-Dichloroethane	6.383	63	436607	48.855	ug/l	99
25) 2-Butanone	7.337	43	800554	234.558	ug/l	94
26) 2,2-Dichloropropane	7.324	77	397092	50.844	ug/l	94
27) cis-1,2-Dichloroethene	7.327	96	281864	49.570	ug/l	87
28) Bromochloromethane	7.657	49	199168	45.671	ug/l	# 79
29) Tetrahydrofuran	7.686	42	531879	236.974	ug/l	92
30) Chloroform	7.817	83	500455	49.854	ug/l	100
31) Cyclohexane	8.094	56	414485	45.254	ug/l	95
32) 1,1,1-Trichloroethane	8.016	97	441211	50.688	ug/l	98
36) 1,1-Dichloropropene	8.220	75	368265	51.408	ug/l	96
37) Ethyl Acetate	7.415	43	340716	49.088	ug/l	100
38) Carbon Tetrachloride	8.206	117	378447	53.558	ug/l	96
39) Methylcyclohexane	9.461	83	414935	52.910	ug/l	92
40) Benzene	8.458	78	1105356	51.549	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071321\
 Data File : VN067717.D
 Acq On : 13 Jul 2021 15:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jul 13 22:17:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.635	41	172501	48.356	ug/l	93
42) 1,2-Dichloroethane	8.531	62	405844	50.966	ug/l	96
43) Isopropyl Acetate	8.563	43	613952	50.285	ug/l	96
44) Trichloroethene	9.217	130	284129	52.211	ug/l	87
45) 1,2-Dichloropropane	9.491	63	278366	51.274	ug/l	98
46) Dibromomethane	9.577	93	203095	52.722	ug/l	95
47) Bromodichloromethane	9.762	83	398525	52.919	ug/l	98
48) Methyl methacrylate	9.561	41	267080	49.459	ug/l	89
49) 1,4-Dioxane	9.571	88	95670	893.998	ug/l	89
51) 4-Methyl-2-Pentanone	10.333	43	1920903	256.093	ug/l	95
52) Toluene	10.507	92	783733	55.797	ug/l	99
53) t-1,3-Dichloropropene	10.719	75	442352	49.317	ug/l	98
54) cis-1,3-Dichloropropene	10.191	75	461629	55.120	ug/l	92
55) 1,1,2-Trichloroethane	10.899	97	291959	54.238	ug/l	98
56) Ethyl methacrylate	10.762	69	462279	55.743	ug/l	91
57) 1,3-Dichloropropane	11.047	76	488243	52.895	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.043	63	442897	203.655	ug/l	92
59) 2-Hexanone	11.087	43	1366577	261.013	ug/l	94
60) Dibromochloromethane	11.240	129	315468	47.857	ug/l	99
61) 1,2-Dibromoethane	11.347	107	305062	55.510	ug/l	99
64) Tetrachloroethene	10.979	164	287017	52.707	ug/l	97
65) Chlorobenzene	11.771	112	820981	51.816	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.843	131	307211	55.123	ug/l	99
67) Ethyl Benzene	11.846	91	1538910	53.354	ug/l	97
68) m/p-Xylenes	11.956	106	1192374	109.299	ug/l	87
69) o-Xylene	12.283	106	578666	53.951	ug/l	88
70) Styrene	12.296	104	983976	57.350	ug/l	93
71) Bromoform	12.460	173	218518	44.571	ug/l #	99
73) Isopropylbenzene	12.581	105	1533385	48.777	ug/l	97
74) N-amyl acetate	12.390	43	539785	47.330	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.830	83	420222	46.059	ug/l	100
76) 1,2,3-Trichloropropane	12.881	75	534256	67.145	ug/l #	100
77) Bromobenzene	12.862	156	346909	49.860	ug/l	74
78) n-propylbenzene	12.921	91	1774058	49.896	ug/l	96
79) 2-Chlorotoluene	13.010	91	1046603	48.723	ug/l	90
80) 1,3,5-Trimethylbenzene	13.061	105	1280473	50.270	ug/l	93
81) trans-1,4-Dichloro-2-b...	12.626	75	123595	50.081	ug/l	95
82) 4-Chlorotoluene	13.106	91	1064624	50.086	ug/l	91
83) tert-Butylbenzene	13.326	119	1073311	49.409	ug/l	89
84) 1,2,4-Trimethylbenzene	13.369	105	1291328	51.739	ug/l	93
85) sec-Butylbenzene	13.503	105	1532204	50.966	ug/l	96
86) p-Isopropyltoluene	13.619	119	1271655	52.417	ug/l	93
87) 1,3-Dichlorobenzene	13.616	146	662080	51.294	ug/l	95
88) 1,4-Dichlorobenzene	13.696	146	651626	49.309	ug/l	96
89) n-Butylbenzene	13.946	91	1049703	50.445	ug/l	98
90) Hexachloroethane	14.211	117	211952	52.603	ug/l	100
91) 1,2-Dichlorobenzene	13.991	146	630404	50.597	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	14.606	75	72102	45.502	ug/l	67
93) 1,2,4-Trichlorobenzene	15.265	180	280640	43.601	ug/l	98
94) Hexachlorobutadiene	15.373	225	136362	52.085	ug/l	98
95) Naphthalene	15.507	128	752258	40.070	ug/l	100
96) 1,2,3-Trichlorobenzene	15.702	180	260903	42.361	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071321\
Data File : VN067717.D
Acq On : 13 Jul 2021 15:07
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Jul 13 22:17:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 13 22:17:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.088 min Scan# 2274
Delta R.T. 0.002 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:168 Resp: 432214
Ion Ratio Lower Upper
168 100
99 61.5 62.9 94.3#



#2
Dichlorodifluoromethane
Concen: 46.012 ug/l
RT: 2.070 min Scan# 30
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 85 Resp: 169036
Ion Ratio Lower Upper
85 100
87 31.9 16.1 48.2









#3
Chloromethane
Concen: 39.132 ug/l
RT: 2.301 min Scan# 116
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 50 Resp: 227693
Ion Ratio Lower Upper
50 100
52 33.6 25.6 38.4



#4
Vinyl Chloride
Concen: 42.206 ug/l
RT: 2.443 min Scan# 169
Delta R.T. 0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 62 Resp: 370636
Ion Ratio Lower Upper
62 100
64 31.4 24.4 36.6









#5
Bromomethane
Concen: 51.671 ug/l
RT: 2.832 min Scan# 314
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 94 Resp: 307537
Ion Ratio Lower Upper
94 100
96 95.8 73.7 110.5



#6
Chloroethane
Concen: 48.944 ug/l
RT: 3.006 min Scan# 379
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 64 Resp: 291607
Ion Ratio Lower Upper
64 100
66 32.6 24.7 37.1









#7
Trichlorofluoromethane
Concen: 55.173 ug/l
RT: 3.365 min Scan# 513
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:101 Resp: 732080
Ion Ratio Lower Upper
101 100
103 64.1 51.2 76.8



#8
Diethyl Ether
Concen: 36.963 ug/l
RT: 3.819 min Scan# 682
Delta R.T. 0.006 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 74 Resp: 129216
Ion Ratio Lower Upper
74 100
45 91.7 50.2 150.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.769 ug/l
RT: 4.199 min Scan# 824
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:101 Resp: 203361
Ion Ratio Lower Upper
101 100
85 47.6 39.6 59.4
151 71.3 54.1 81.1



#10
Methyl Iodide
Concen: 51.556 ug/l
RT: 4.414 min Scan# 904
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:142 Resp: 246010
Ion Ratio Lower Upper
142 100
127 47.4 43.1 64.7
141 14.4 11.4 17.0









#11
Tert butyl alcohol
Concen: 221.595 ug/l
RT: 5.385 min Scan# 1266
Delta R.T. -0.005 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 59 Resp: 208002
Ion Ratio Lower Upper
59 100
57 5.4 8.7 13.1#



#12
1,1-Dichloroethene
Concen: 48.896 ug/l
RT: 4.170 min Scan# 813
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 96 Resp: 187912
Ion Ratio Lower Upper
96 100
61 171.9 152.6 229.0
98 64.4 50.9 76.3









#13
Acrolein
Concen: 275.637 ug/l
RT: 4.036 min Scan# 763
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 56 Resp: 125934
Ion Ratio Lower Upper
56 100
55 70.6 57.4 86.0



#14
Allyl chloride
Concen: 47.334 ug/l
RT: 4.835 min Scan# 1061
Delta R.T. 0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 41 Resp: 334395
Ion Ratio Lower Upper
41 100
39 81.3 60.9 91.3
76 34.8 24.3 36.5









#15
Acrylonitrile
Concen: 239.885 ug/l
RT: 5.554 min Scan# 1329
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 53 Resp: 558490
Ion Ratio Lower Upper
53 100
52 82.5 66.2 99.2
51 37.8 30.2 45.4



#16
Acetone
Concen: 220.466 ug/l
RT: 4.285 min Scan# 856
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 43 Resp: 453140
Ion Ratio Lower Upper
43 100
58 28.9 21.2 31.8









#17
Carbon Disulfide
Concen: 46.869 ug/l
RT: 4.521 min Scan# 944
Delta R.T. 0.002 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 498948
Ion Ratio Lower Upper
76 100
78 8.8 7.7 11.5



#18
Methyl Acetate
Concen: 43.748 ug/l
RT: 4.854 min Scan# 1068
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 43 Resp: 252560
Ion Ratio Lower Upper
43 100
74 24.8 17.4 26.0









#19
Methyl tert-butyl Ether
Concen: 48.877 ug/l
RT: 5.618 min Scan# 1353
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 73 Resp: 778399
Ion Ratio Lower Upper
73 100
57 20.7 17.0 25.6



#20
Methylene Chloride
Concen: 49.426 ug/l
RT: 5.093 min Scan# 1157
Delta R.T. 0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 84 Resp: 237196
Ion Ratio Lower Upper
84 100
49 122.7 110.9 166.3
51 37.9 32.6 49.0
86 66.5 51.6 77.4









#21
trans-1,2-Dichloroethene
Concen: 50.111 ug/l
RT: 5.589 min Scan# 1342
Delta R.T. -0.002 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 96 Resp: 227690
Ion Ratio Lower Upper
96 100
61 133.0 123.5 185.3
98 60.7 48.6 72.8



#22
Diisopropyl ether
Concen: 50.562 ug/l
RT: 6.498 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 45 Resp: 810560
Ion Ratio Lower Upper
45 100
43 60.9 49.4 74.2
87 29.1 20.9 31.3
59 12.3 9.0 13.4









#23
 Vinyl Acetate
 Concen: 248.479 ug/l
 RT: 6.434 min Scan# 1657
 Delta R.T. -0.000 min
 Lab File: VN067717.D
 Acq: 13 Jul 2021 15:07




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion: 43 Resp: 3405509
 Ion Ratio Lower Upper
 43 100
 86 10.5 7.3 10.9



#24
 1,1-Dichloroethane
 Concen: 48.855 ug/l
 RT: 6.383 min Scan# 1638
 Delta R.T. -0.002 min
 Lab File: VN067717.D
 Acq: 13 Jul 2021 15:07

Tgt Ion: 63 Resp: 436607
 Ion Ratio Lower Upper
 63 100
 98 6.3 3.0 9.0
 100 4.0 1.9 5.7





#25

2-Butanone

Concen: 234.558 ug/l

RT: 7.337 min Scan# 1994

Delta R.T. 0.000 min

Lab File: VN067717.D





Acq: 13 Jul 2021 15:07

Tgt Ion: 43 Resp: 800554

Ion Ratio Lower Upper

43 100

72 27.4 19.7 29.5



#26

2,2-Dichloropropane

Concen: 50.844 ug/l

RT: 7.324 min Scan# 1989

Delta R.T. -0.000 min

Lab File: VN067717.D




Acq: 13 Jul 2021 15:07


Tgt Ion: 77 Resp: 397092

Ion Ratio Lower Upper

77 100

97 22.2 9.8 29.3









#27
 cis-1,2-Dichloroethene
 Concen: 49.570 ug/l
 RT: 7.327 min Scan# 1990
 Delta R.T. 0.003 min
 Lab File: VN067717.D
 Acq: 13 Jul 2021 15:07




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion: 96 Resp: 281864
 Ion Ratio Lower Upper
 96 100
 61 144.9 0.0 338.8
 98 64.3 0.0 129.8



#28
 Bromochloromethane
 Concen: 45.671 ug/l
 RT: 7.657 min Scan# 2113
 Delta R.T. -0.000 min
 Lab File: VN067717.D
 Acq: 13 Jul 2021 15:07


Tgt Ion: 49 Resp: 199168
 Ion Ratio Lower Upper
 49 100
 129 2.3 0.0 3.8
 130 78.7 49.7 74.5#








#29
Tetrahydrofuran
Concen: 236.974 ug/l
RT: 7.686 min Scan# 2124
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 42 Resp: 531879
Ion Ratio Lower Upper
42 100
72 46.0 32.8 49.2
71 44.1 31.2 46.8



#30
Chloroform
Concen: 49.854 ug/l
RT: 7.817 min Scan# 2173
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 83 Resp: 500455
Ion Ratio Lower Upper
83 100
85 64.4 51.5 77.3









#31
Cyclohexane
Concen: 45.254 ug/l
RT: 8.094 min Scan# 2276
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 56 Resp: 414485
Ion Ratio Lower Upper
56 100
69 33.9 27.3 40.9
84 87.9 64.9 97.3



#32
1,1,1-Trichloroethane
Concen: 50.688 ug/l
RT: 8.016 min Scan# 2247
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 97 Resp: 441211
Ion Ratio Lower Upper
97 100
99 63.8 50.2 75.2
61 45.2 38.2 57.2









#33

1,2-Dichloroethane-d4
Concen: 47.958 ug/l
RT: 8.437 min Scan# 2404
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Instrument :
MSVOA_N
ClientSampled :




Tgt Ion: 65 Resp: 313267
Ion Ratio Lower Upper
65 100
67 50.7 0.0 98.6




#34

1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.968 min Scan# 2602
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 114 Resp: 771182
Ion Ratio Lower Upper
114 100
63 21.3 0.0 49.0
88 15.5 0.0 36.6





#35

Dibromofluoromethane

Concen: 49.804 ug/l

RT: 8.021 min Scan# 2249

Delta R.T. 0.000 min

Lab File: VN067717.D

Acq: 13 Jul 2021 15:07





Tgt Ion:113 Resp: 232050

Ion Ratio Lower Upper

113 100

111 103.2 83.8 125.8

192 16.2 13.4 20.0



#36

1,1-Dichloropropene

Concen: 51.408 ug/l

RT: 8.220 min Scan# 2323

Delta R.T. -0.002 min

Lab File: VN067717.D

Acq: 13 Jul 2021 15:07




Tgt Ion: 75 Resp: 368265


Ion Ratio Lower Upper

75 100

110 35.5 15.7 47.0

77 31.3 24.6 37.0









#37
Ethyl Acetate
Concen: 49.088 ug/l
RT: 7.415 min Scan# 2023
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 340716
Ion Ratio Lower Upper
43 100
61 13.6 10.6 16.0
70 10.1 8.1 12.1



#38
Carbon Tetrachloride
Concen: 53.558 ug/l
RT: 8.206 min Scan# 2318
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 117 Resp: 378447
Ion Ratio Lower Upper
117 100
119 97.9 75.4 113.2
121 31.2 23.5 35.3









#39
Methylcyclohexane
Concen: 52.910 ug/l
RT: 9.461 min Scan# 2786
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 83 Resp: 414935
Ion Ratio Lower Upper
83 100
55 74.8 66.6 99.8
98 47.7 35.3 52.9



#40
Benzene
Concen: 51.549 ug/l
RT: 8.458 min Scan# 2412
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 78 Resp: 1105356
Ion Ratio Lower Upper
78 100
77 23.7 18.9 28.3









#41
Methacrylonitrile
Concen: 48.356 ug/l
RT: 7.635 min Scan# 2105
Delta R.T. -0.005 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 41 Resp: 172501




Ion	Ratio	Lower	Upper
41	100		
39	61.0	47.6	71.4
67	81.6	57.7	86.5
52	35.1	26.2	39.4




#42
1,2-Dichloroethane
Concen: 50.966 ug/l
RT: 8.531 min Scan# 2439
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 62 Resp: 405844

Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	14.2









#43
Isopropyl Acetate
Concen: 50.285 ug/l
RT: 8.563 min Scan# 2451
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 613952
Ion Ratio Lower Upper
43 100
61 25.7 19.0 28.4
87 13.8 9.5 14.3



#44
Trichloroethene
Concen: 52.211 ug/l
RT: 9.217 min Scan# 2695
Delta R.T. 0.002 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:130 Resp: 284129
Ion Ratio Lower Upper
130 100
95 102.9 0.0 234.4









#45
1,2-Dichloropropane
Concen: 51.274 ug/l
RT: 9.491 min Scan# 2797
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 63 Resp: 278366
Ion Ratio Lower Upper
63 100
65 31.6 24.3 36.5



#46
Dibromomethane
Concen: 52.722 ug/l
RT: 9.577 min Scan# 2829
Delta R.T. -0.002 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 93 Resp: 203095
Ion Ratio Lower Upper
93 100
95 82.7 66.1 99.1
174 84.6 61.0 91.4









#47
Bromodichloromethane
Concen: 52.919 ug/l
RT: 9.762 min Scan# 2898
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 398525
Ion Ratio Lower Upper
83 100
85 63.6 49.9 74.9
127 8.1 6.1 9.1



#48
Methyl methacrylate
Concen: 49.459 ug/l
RT: 9.561 min Scan# 2823
Delta R.T. -0.002 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 41 Resp: 267080
Ion Ratio Lower Upper
41 100
69 92.1 64.6 96.8
39 74.2 53.3 79.9





#49

1,4-Dioxane

Concen: 893.998 ug/l

RT: 9.571 min Scan# 2827

Delta R.T. 0.000 min

Lab File: VN067717.D

Acq: 13 Jul 2021 15:07

Tgt Ion: 88 Resp: 95670

Ion Ratio Lower Upper

88 100





43 31.1 28.8 43.2

58 65.4 60.4 90.6

Instrument :

MSVOA_N

ClientSampleId :



#50

Toluene-d8

Concen: 51.165 ug/l

RT: 10.443 min Scan# 3152

Delta R.T. 0.000 min

Lab File: VN067717.D




Acq: 13 Jul 2021 15:07


Tgt Ion: 98 Resp: 997862

Ion Ratio Lower Upper

98 100

100 63.8 51.7 77.5









#51
4-Methyl-2-Pentanone
Concen: 256.093 ug/l
RT: 10.333 min Scan# 3111
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 1920903
Ion Ratio Lower Upper
43 100
58 38.5 28.6 43.0



#52
Toluene
Concen: 55.797 ug/l
RT: 10.507 min Scan# 3176
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 92 Resp: 783733
Ion Ratio Lower Upper
92 100
91 176.1 139.6 209.4









#53
t-1,3-Dichloropropene
Concen: 49.317 ug/l
RT: 10.719 min Scan# 3255
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 75 Resp: 442352
Ion Ratio Lower Upper
75 100
77 30.4 25.1 37.7



#54
cis-1,3-Dichloropropene
Concen: 55.120 ug/l
RT: 10.191 min Scan# 3058
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 75 Resp: 461629
Ion Ratio Lower Upper
75 100
77 30.9 24.6 37.0
39 54.3 50.7 76.1









#55
1,1,2-Trichloroethane
Concen: 54.238 ug/l
RT: 10.899 min Scan# 3322
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	97	Resp:	291959
Ion	Ratio	Lower	Upper
97	100		
83	87.0	72.0	108.0
85	56.4	46.8	70.2
99	63.1	49.8	74.6



#56
Ethyl methacrylate
Concen: 55.743 ug/l
RT: 10.762 min Scan# 3271
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:	69	Resp:	462279
Ion	Ratio	Lower	Upper
69	100		
41	65.1	58.6	88.0
39	43.1	38.3	57.5









#57
1,3-Dichloropropane
Concen: 52.895 ug/l
RT: 11.047 min Scan# 3377
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 488243
Ion Ratio Lower Upper
76 100
78 32.1 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 203.655 ug/l
RT: 10.043 min Scan# 3003
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 63 Resp: 442897
Ion Ratio Lower Upper
63 100
106 28.6 19.8 29.8









#59
2-Hexanone
Concen: 261.013 ug/l
RT: 11.087 min Scan# 3392
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 1366577
Ion Ratio Lower Upper
43 100
58 53.6 24.7 74.1



#60
Dibromochloromethane
Concen: 47.857 ug/l
RT: 11.240 min Scan# 3449
Delta R.T. -0.002 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:129 Resp: 315468
Ion Ratio Lower Upper
129 100
127 77.9 38.6 115.8









#61
1,2-Dibromoethane
Concen: 55.510 ug/l
RT: 11.347 min Scan# 3489
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 305062
Ion Ratio Lower Upper
107 100
109 94.5 75.0 112.4



#62
4-Bromofluorobenzene
Concen: 53.212 ug/l
RT: 12.733 min Scan# 4006
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 95 Resp: 366601
Ion Ratio Lower Upper
95 100
174 75.9 0.0 122.0
176 72.9 0.0 117.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.747 min Scan# 3638
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117 Resp: 771438
Ion Ratio Lower Upper
117 100
82 55.0 53.1 79.7
119 32.5 25.5 38.3



#64
Tetrachloroethene
Concen: 52.707 ug/l
RT: 10.979 min Scan# 3352
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:164 Resp: 287017
Ion Ratio Lower Upper
164 100
166 127.4 102.1 153.1
129 105.2 86.9 130.3
131 98.3 83.6 125.4









#65
Chlorobenzene
Concen: 51.816 ug/l
RT: 11.771 min Scan# 3647
Delta R.T. -0.002 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:112 Resp: 820981
Ion Ratio Lower Upper
112 100
114 31.6 24.6 37.0



#66
1,1,1,2-Tetrachloroethane
Concen: 55.123 ug/l
RT: 11.843 min Scan# 3674
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:131 Resp: 307211
Ion Ratio Lower Upper
131 100
133 96.2 47.5 142.5
119 65.8 33.1 99.3









#67
Ethyl Benzene
Concen: 53.354 ug/l
RT: 11.846 min Scan# 3675
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 1538910
Ion Ratio Lower Upper
91 100
106 30.7 23.4 35.0



#68
m/p-Xylenes
Concen: 109.299 ug/l
RT: 11.956 min Scan# 3716
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:106 Resp: 1192374
Ion Ratio Lower Upper
106 100
91 204.6 180.6 270.8









#69
o-Xylene
Concen: 53.951 ug/l
RT: 12.283 min Scan# 3838
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 578666
Ion Ratio Lower Upper
106 100
91 217.1 118.3 354.9



#70
Styrene
Concen: 57.350 ug/l
RT: 12.296 min Scan# 3843
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:104 Resp: 983976
Ion Ratio Lower Upper
104 100
78 50.1 48.2 72.2
103 55.3 44.5 66.7









#71
Bromoform
Concen: 44.571 ug/l
RT: 12.460 min Scan# 3904
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 218518
Ion Ratio Lower Upper
173 100
175 48.3 23.9 71.8
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.675 min Scan# 4357
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:152 Resp: 389621
Ion Ratio Lower Upper
152 100
115 60.2 36.1 108.5
150 173.6 0.0 351.8









#73
Isopropylbenzene
Concen: 48.777 ug/l
RT: 12.581 min Scan# 3949
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:105 Resp: 1533385
Ion Ratio Lower Upper
105 100
120 26.3 12.5 37.5



#74
N-ethyl acetate
Concen: 47.330 ug/l
RT: 12.390 min Scan# 3878
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 43 Resp: 539785
Ion Ratio Lower Upper
43 100
70 41.2 29.4 44.2
55 23.4 17.0 25.4
61 23.5 17.6 26.4









#75
1,1,2,2-Tetrachloroethane
Concen: 46.059 ug/l
RT: 12.830 min Scan# 4042
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 83 Resp: 420222
Ion Ratio Lower Upper
83 100
131 10.0 4.8 14.4
85 64.3 32.2 96.6



#76
1,2,3-Trichloropropane
Concen: 67.145 ug/l
RT: 12.881 min Scan# 4061
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 75 Resp: 534256
Ion Ratio Lower Upper
75 100
77 134.8 0.0 0.0#









#77
Bromobenzene
Concen: 49.860 ug/l
RT: 12.862 min Scan# 4054
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 346909
Ion Ratio Lower Upper
156 100
77 207.5 136.2 408.6
158 96.2 48.4 145.3



#78
n-propylbenzene
Concen: 49.896 ug/l
RT: 12.921 min Scan# 4076
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 91 Resp: 1774058
Ion Ratio Lower Upper
91 100
120 23.0 10.4 31.4









#79
2-Chlorotoluene
Concen: 48.723 ug/l
RT: 13.010 min Scan# 4109
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 1046603
Ion Ratio Lower Upper
91 100
126 34.5 14.5 43.5



#80
1,3,5-Trimethylbenzene
Concen: 50.270 ug/l
RT: 13.061 min Scan# 4128
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 105 Resp: 1280473
Ion Ratio Lower Upper
105 100
120 48.9 22.2 66.6









#81
trans-1,4-Dichloro-2-butene
Concen: 50.081 ug/l
RT: 12.626 min Scan# 3966
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 123595
Ion Ratio Lower Upper
75 100
53 101.9 87.6 131.4
89 47.9 37.8 56.8



#82
4-Chlorotoluene
Concen: 50.086 ug/l
RT: 13.106 min Scan# 4145
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07


Tgt Ion: 91 Resp: 1064624
Ion Ratio Lower Upper
91 100
126 34.3 14.8 44.4








#83
tert-Butylbenzene
Concen: 49.409 ug/l
RT: 13.326 min Scan# 4227
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:119 Resp: 1073311
Ion Ratio Lower Upper
119 100
91 66.8 39.5 118.5
134 25.7 12.0 36.0




#84
1,2,4-Trimethylbenzene
Concen: 51.739 ug/l
RT: 13.369 min Scan# 4243
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07



Tgt Ion:105 Resp: 1291328
Ion Ratio Lower Upper
105 100
120 45.8 20.6 61.8









#85
 sec-Butylbenzene
 Concen: 50.966 ug/l
 RT: 13.503 min Scan# 4293
 Delta R.T. 0.000 min
 Lab File: VN067717.D
 Acq: 13 Jul 2021 15:07




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion:105 Resp: 1532204
 Ion Ratio Lower Upper
 105 100
 134 20.1 9.0 27.1



#86
 p-Isopropyltoluene
 Concen: 52.417 ug/l
 RT: 13.619 min Scan# 4336
 Delta R.T. -0.000 min
 Lab File: VN067717.D
 Acq: 13 Jul 2021 15:07

Tgt Ion:119 Resp: 1271655
 Ion Ratio Lower Upper
 119 100
 134 26.5 12.4 37.4
 91 24.1 14.8 44.3









#87

1,3-Dichlorobenzene
Concen: 51.294 ug/l
RT: 13.616 min Scan# 4335
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Instrument :
MSVOA_N
ClientSampleId :




Tgt Ion:146 Resp: 662080
Ion Ratio Lower Upper
146 100
111 41.2 24.3 72.8
148 63.8 32.0 96.0




#88

1,4-Dichlorobenzene
Concen: 49.309 ug/l
RT: 13.696 min Scan# 4365
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:146 Resp: 651626
Ion Ratio Lower Upper
146 100
111 40.5 23.4 70.2
148 63.8 31.8 95.3









#89
n-Butylbenzene
Concen: 50.445 ug/l
RT: 13.946 min Scan# 4458
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 91 Resp: 1049703




Ion	Ratio	Lower	Upper
91	100		
92	51.2	25.7	77.1
134	25.8	11.2	33.6




#90
Hexachloroethane
Concen: 52.603 ug/l
RT: 14.211 min Scan# 4557
Delta R.T. -0.003 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion: 117 Resp: 211952

Ion	Ratio	Lower	Upper
117	100		
201	66.7	33.5	100.5









#91
1,2-Dichlorobenzene
Concen: 50.597 ug/l
RT: 13.991 min Scan# 4475
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion	Ratio	Lower	Upper
146	100		
111	42.2	24.9	74.6
148	63.3	31.6	95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 45.502 ug/l
RT: 14.606 min Scan# 4704
Delta R.T. 0.001 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07


Tgt Ion	Ratio	Lower	Upper
75	100		
155	81.4	29.8	89.5
157	106.1	37.6	112.8








#93
1,2,4-Trichlorobenzene
Concen: 43.601 ug/l
RT: 15.265 min Scan# 4950
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07


Instrument :
MSVOA_N
ClientSampled :





Tgt Ion:180 Resp: 280640
Ion Ratio Lower Upper
180 100
182 93.7 47.5 142.5
145 32.1 17.2 51.6




#94
Hexachlorobutadiene
Concen: 52.085 ug/l
RT: 15.373 min Scan# 4990
Delta R.T. -0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07



Tgt Ion:225 Resp: 136362
Ion Ratio Lower Upper
225 100
223 65.5 31.4 94.0
227 62.8 31.2 93.6









#95
Naphthalene
Concen: 40.070 ug/l
RT: 15.507 min Scan# 5040
Delta R.T. -0.002 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Instrument :
MSVOA_N
ClientSampled :

Tgt Ion:128	Resp:	752258
Ion Ratio	Lower	Upper
128	100	
127	13.2	10.6 15.8
129	11.0	8.6 12.8



#96
1,2,3-Trichlorobenzene
Concen: 42.361 ug/l
RT: 15.702 min Scan# 5113
Delta R.T. 0.000 min
Lab File: VN067717.D
Acq: 13 Jul 2021 15:07

Tgt Ion:180	Resp:	260903
Ion Ratio	Lower	Upper
180	100	
182	94.6	47.8 143.3
145	34.1	18.2 54.6

