

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071321\
 Data File : VN067713.D
 Acq On : 13 Jul 2021 13:11
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 13 22:16:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	354347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	651991	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	612193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	222356	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	272135	50.816	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.640%	
35) Dibromofluoromethane	8.024	113	198212	50.318	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.640%	
50) Toluene-d8	10.443	98	827646	50.196	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.400%	
62) 4-Bromofluorobenzene	12.733	95	267230	45.879	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 91.760%	
Target Compounds						
						Qvalue
16) Acetone	4.299	43	7275	4.317	ug/l	91
30) Chloroform	7.817	83	29269	3.556	ug/l	98
95) Naphthalene	15.512	128	19875	4.728	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071321\
Data File : VN067713.D
Acq On : 13 Jul 2021 13:11
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 13 22:16:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

