

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071322\
 Data File : VN073453.D
 Acq On : 13 Jul 2022 17:04
 Operator : JC\MD
 Sample : N3707-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 N48967

Quant Time: Jul 14 07:04:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	301360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	501125	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	469840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	227659	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	193272	44.426	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.860%
35) Dibromofluoromethane	7.951	113	161937	48.571	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.140%
50) Toluene-d8	10.380	98	531337	44.871	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.740%
62) 4-Bromofluorobenzene	12.668	95	234039	51.430	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.860%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	13279	5.438	ug/l	90
25) 2-Butanone	7.275	43	15516	3.991	ug/l	90
27) cis-1,2-Dichloroethene	7.239	96	5261	1.172	ug/l	85
29) Tetrahydrofuran	7.622	42	14576	5.475	ug/l #	83
44) Trichloroethene	9.151	130	2142	0.532	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071322\
Data File : VN073453.D
Acq On : 13 Jul 2022 17:04
Operator : JC\MD
Sample : N3707-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
N48967

Quant Time: Jul 14 07:04:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

