

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071323\
 Data File : VN078496.D
 Acq On : 13 Jul 2023 14:11
 Operator : JC\MD
 Sample : 03610-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230712039-08-TRIP-BLANK

Quant Time: Jul 14 01:33:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	409994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	744163	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	694817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	254570	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	279872	50.919	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.840%
35) Dibromofluoromethane	8.177	113	275533	54.544	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.080%
50) Toluene-d8	10.576	98	873254	45.760	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.520%
62) 4-Bromofluorobenzene	12.853	95	315093	53.604	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071323\
Data File : VN078496.D
Acq On : 13 Jul 2023 14:11
Operator : JC\MD
Sample : 03610-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230712039-08-TRIP-BLANK

Quant Time: Jul 14 01:33:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

