

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071323\
 Data File : VN078497.D
 Acq On : 13 Jul 2023 14:35
 Operator : JC\MD
 Sample : 03611-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-RW5A-20230712

Quant Time: Jul 14 01:33:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	404884	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	736695	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	675513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	212893	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	265451	48.905	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.800%		
35) Dibromofluoromethane	8.177	113	278422	55.674	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	111.340%		
50) Toluene-d8	10.576	98	856644	45.344	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.680%		
62) 4-Bromofluorobenzene	12.859	95	283465	48.713	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.420%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.136	85	3540	0.713	ug/l	82
9) 1,1,2-Trichlorotrifluo...	4.383	101	83016	18.796	ug/l	96
12) 1,1-Dichloroethene	4.353	96	35467	8.302	ug/l	95
27) cis-1,2-Dichloroethene	7.494	96	11434	2.053	ug/l	82
30) Chloroform	7.971	83	6896	0.752	ug/l	95
44) Trichloroethene	9.359	130	1534628	266.166	ug/l	97
55) 1,1,2-Trichloroethane	11.024	97	3512	0.672	ug/l #	85
64) Tetrachloroethene	11.112	164	5734	1.129	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071323\
Data File : VN078497.D
Acq On : 13 Jul 2023 14:35
Operator : JC\MD
Sample : 03611-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-RW5A-20230712

Quant Time: Jul 14 01:33:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

