

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071421\
 Data File : VN067746.D
 Acq On : 14 Jul 2021 16:25
 Operator : JC/MD
 Sample : M2960-01RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JULY)RE

Manual Integrations
 APPROVED

MMDadoda
 7/16/2021 10:43:24 AM

Quant Time: Jul 14 17:16:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N071421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 14 01:19:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.659	128	58011	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	336196	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.747	117	301474	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	155700	33.222	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	110.733%#
60) 4-Bromofluorobenzene	12.734	95	122225	24.964	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.200%
63) Toluene-d8	10.446	98	414481	29.891	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.633%
Target Compounds						
15) Acetone	4.307	58	2862m	6.836	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071421\
Data File : VN067746.D
Acq On : 14 Jul 2021 16:25
Operator : JC/MD
Sample : M2960-01RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(JULY)RE

Manual Integrations
APPROVED

MMDadoda
7/16/2021 10:43:24 AM

Quant Time: Jul 14 17:16:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N071421W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 14 01:19:19 2021
Response via : Initial Calibration

