

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071421\
 Data File : VN067748.D
 Acq On : 14 Jul 2021 17:08
 Operator : JC/MD
 Sample : M3006-01
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-070921

Quant Time: Jul 14 17:44:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	310965	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.965	114	580890	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	526364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	201757	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	241991	51.491	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.980%
35) Dibromofluoromethane	8.016	113	168429	47.991	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.980%
50) Toluene-d8	10.443	98	725990	49.420	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.733	95	235217	45.326	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.660%
Target Compounds						
20) Methylene Chloride	5.058	84	9196	1.996	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071421\
Data File : VN067748.D
Acq On : 14 Jul 2021 17:08
Operator : JC/MD
Sample : M3006-01
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-02-070921

Quant Time: Jul 14 17:44:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

