

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071421\
 Data File : VN067734.D
 Acq On : 14 Jul 2021 10:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 14 17:44:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.425	0.458	-7.8	103	0.00
3 P	Chloromethane	0.673	0.576	14.4	90	0.00
4 C	Vinyl Chloride	1.016	0.877	13.7#	89	0.00
5 T	Bromomethane	0.838	0.679	19.0	88	0.00
6 T	Chloroethane	0.753	0.646	14.2	89	0.00
7 T	Trichlorofluoromethane	1.885	1.632	13.4	92	0.00
8 T	Diethyl Ether	0.404	0.320	20.8	68	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.463	0.485	-4.8	105	0.00
10 T	Methyl Iodide	0.552	0.599	-8.5	98	0.00
11 T	Tert butyl alcohol	0.109	0.105	3.7	96	0.00
12 CM	1,1-Dichloroethene	0.445	0.448	-0.7#	100	0.00
13 T	Acrolein	0.058	0.069	-19.0	126	0.00
14 T	Allyl chloride	0.817	0.804	1.6	96	0.00
15 T	Acrylonitrile	0.269	0.280	-4.1	96	0.00
16 T	Acetone	0.238	0.241	-1.3	107	0.00
17 T	Carbon Disulfide	1.232	1.211	1.7	95	0.00
18 T	Methyl Acetate	0.668	0.650	2.7	99	0.00
19 T	Methyl tert-butyl Ether	1.842	1.906	-3.5	96	0.00
20 T	Methylene Chloride	0.629	0.574	8.7	98	0.00
21 T	trans-1,2-Dichloroethene	0.526	0.541	-2.9	96	0.00
22 T	Diisopropyl ether	1.855	1.920	-3.5	94	0.00
23 T	Vinyl Acetate	1.585	1.626	-2.6	93	0.00
24 P	1,1-Dichloroethane	1.034	1.027	0.7	94	0.00
25 T	2-Butanone	0.395	0.398	-0.8	95	0.00
26 T	2,2-Dichloropropane	0.903	0.936	-3.7	97	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.665	-1.1	93	0.00
28 T	Bromochloromethane	0.504	0.472	6.3	93	0.00
29 T	Tetrahydrofuran	0.260	0.260	0.0	91	0.00
30 C	Chloroform	1.161	1.157	0.3#	93	0.00
31 T	Cyclohexane	1.060	0.933	12.0	90	0.00
32 T	1,1,1-Trichloroethane	1.007	1.013	-0.6	91	0.00
33 S	1,2-Dichloroethane-d4	0.756	0.728	3.7	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.302	0.316	-4.6	97	0.00
36 T	1,1-Dichloropropene	0.464	0.478	-3.0	91	0.00
37 T	Ethyl Acetate	0.450	0.463	-2.9	92	0.00
38 T	Carbon Tetrachloride	0.458	0.495	-8.1	91	0.00
39 T	Methylcyclohexane	0.508	0.542	-6.7	93	0.00
40 TM	Benzene	1.390	1.427	-2.7	90	0.00
41 T	Methacrylonitrile	0.231	0.239	-3.5	91	0.00
42 TM	1,2-Dichloroethane	0.516	0.534	-3.5	93	0.00
43 T	Isopropyl Acetate	0.792	0.820	-3.5	92	0.00
44 TM	Trichloroethene	0.353	0.370	-4.8	94	0.00
45 C	1,2-Dichloropropane	0.352	0.359	-2.0#	90	0.00
46 T	Dibromomethane	0.250	0.267	-6.8	93	0.00
47 T	Bromodichloromethane	0.488	0.518	-6.1	91	0.00
48 T	Methyl methacrylate	0.350	0.357	-2.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071421\
 Data File : VN067734.D
 Acq On : 14 Jul 2021 10:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 14 17:44:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	88	0.00
50 S	Toluene-d8	1.264	1.320	-4.4	94	0.00
51 T	4-Methyl-2-Pentanone	0.486	0.516	-6.2	92	0.00
52 CM	Toluene	0.911	0.986	-8.2#	90	0.00
53 T	t-1,3-Dichloropropene	0.501	0.584	-16.6	93	0.00
54 T	cis-1,3-Dichloropropene	0.543	0.600	-10.5	92	0.00
55 T	1,1,2-Trichloroethane	0.349	0.382	-9.5	91	0.00
56 T	Ethyl methacrylate	0.538	0.608	-13.0	91	0.00
57 T	1,3-Dichloropropane	0.598	0.632	-5.7	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.111	0.120	-8.1	82	0.00
59 T	2-Hexanone	0.339	0.373	-10.0	92	0.00
60 T	Dibromochloromethane	0.354	0.411	-16.1	90	0.00
61 T	1,2-Dibromoethane	0.356	0.399	-12.1	91	0.00
62 S	4-Bromofluorobenzene	0.447	0.479	-7.2	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.353	0.374	-5.9	94	0.00
65 PM	Chlorobenzene	1.027	1.091	-6.2	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.410	-13.6	88	0.00
67 C	Ethyl Benzene	1.869	2.008	-7.4#	88	0.00
68 T	m/p-Xylenes	0.707	0.781	-10.5	88	0.00
69 T	o-Xylene	0.695	0.761	-9.5	88	0.00
70 T	Styrene	1.112	1.288	-15.8	88	0.00
71 P	Bromoform	0.244	0.297	-21.7	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	4.034	4.022	0.3	89	0.00
74 T	N-amyl acetate	1.464	1.504	-2.7	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.171	1.158	1.1	88	0.00
76 T	1,2,3-Trichloropropane	1.021	0.998	2.3	96	0.00
77 T	Bromobenzene	0.893	0.923	-3.4	84	0.00
78 T	n-propylbenzene	4.563	4.642	-1.7	90	0.00
79 T	2-Chlorotoluene	2.757	2.752	0.2	89	0.00
80 T	1,3,5-Trimethylbenzene	3.269	3.369	-3.1	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.317	0.339	-6.9	88	0.00
82 T	4-Chlorotoluene	2.728	2.791	-2.3	90	0.00
83 T	tert-Butylbenzene	2.788	2.814	-0.9	86	0.00
84 T	1,2,4-Trimethylbenzene	3.203	3.340	-4.3	89	0.00
85 T	sec-Butylbenzene	3.858	3.995	-3.6	91	0.00
86 T	p-Isopropyltoluene	3.113	3.345	-7.5	90	0.00
87 T	1,3-Dichlorobenzene	1.656	1.787	-7.9	88	0.00
88 T	1,4-Dichlorobenzene	1.696	1.747	-3.0	87	0.00
89 T	n-Butylbenzene	2.670	2.816	-5.5	92	0.00
90 T	Hexachloroethane	0.517	0.552	-6.8	86	0.00
91 T	1,2-Dichlorobenzene	1.599	1.692	-5.8	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.203	0.209	-3.0	86	0.00
93 T	1,2,4-Trichlorobenzene	0.688	0.780	-13.4	85	0.00
94 T	Hexachlorobutadiene	0.336	0.375	-11.6	88	0.00
95 T	Naphthalene	1.940	2.254	-16.2	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071421\
Data File : VN067734.D
Acq On : 14 Jul 2021 10:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jul 14 17:44:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.665	0.740	-11.3	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6