

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071422\
 Data File : VN073458.D
 Acq On : 14 Jul 2022 12:59
 Operator : JC\MD
 Sample : VN0714WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0714WBL01

Quant Time: Jul 15 02:00:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	271758	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	462806	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	421733	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	206175	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	199767	50.920	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.840%
35) Dibromofluoromethane	7.951	113	161842	52.561	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.120%
50) Toluene-d8	10.381	98	524121	47.926	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.860%
62) 4-Bromofluorobenzene	12.669	95	219675	52.271	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071422\
Data File : VN073458.D
Acq On : 14 Jul 2022 12:59
Operator : JC\MD
Sample : VN0714WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0714WBL01

Quant Time: Jul 15 02:00:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

