

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071422\
 Data File : VN073466.D
 Acq On : 14 Jul 2022 17:28
 Operator : JC\MD
 Sample : N3715-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB-01-071322

Quant Time: Jul 15 02:02:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	251437	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.904	114	436605	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	392738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	183627	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	171419	47.226	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.460%
35) Dibromofluoromethane	7.957	113	143387	49.362	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.720%
50) Toluene-d8	10.380	98	453799	43.986	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.980%
62) 4-Bromofluorobenzene	12.669	95	193351	48.768	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.540%
Target Compounds						
16) Acetone	4.246	43	4315	2.118	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071422\
Data File : VN073466.D
Acq On : 14 Jul 2022 17:28
Operator : JC\MD
Sample : N3715-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB-01-071322

Quant Time: Jul 15 02:02:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

