

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071422\
 Data File : VN073471.D
 Acq On : 14 Jul 2022 19:31
 Operator : JC\MD
 Sample : N3663-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VB16181

Quant Time: Jul 15 02:03:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	254537	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	440852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	412354	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	195236	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	178293	48.521	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.040%
35) Dibromofluoromethane	7.951	113	144936	49.415	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.820%
50) Toluene-d8	10.380	98	470351	45.151	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.300%
62) 4-Bromofluorobenzene	12.668	95	200254	50.022	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.040%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	144339	69.980	ug/l	94
20) Methylene Chloride	5.010	84	24404	6.733	ug/l	90
25) 2-Butanone	7.263	43	45685	13.912	ug/l #	87
37) Ethyl Acetate	7.345	43	20733	3.040	ug/l #	95
43) Isopropyl Acetate	8.498	43	67170	6.854	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071422\
Data File : VN073471.D
Acq On : 14 Jul 2022 19:31
Operator : JC\MD
Sample : N3663-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VB16181

Quant Time: Jul 15 02:03:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

