

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071423\
 Data File : VN078514.D
 Acq On : 14 Jul 2023 11:56
 Operator : JC\MD
 Sample : 03624-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 15 03:59:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	399967	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	728085	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	686120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	221700	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	278847	52.004	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.000%
35) Dibromofluoromethane	8.177	113	277414	56.129	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.260%
50) Toluene-d8	10.576	98	861610	46.147	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.300%
62) 4-Bromofluorobenzene	12.853	95	294794	51.259	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071423\
Data File : VN078514.D
Acq On : 14 Jul 2023 11:56
Operator : JC\MD
Sample : 03624-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 15 03:59:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

