

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071423\
 Data File : VN078530.D
 Acq On : 14 Jul 2023 18:27
 Operator : JC\MD
 Sample : 03637-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 A508

Quant Time: Jul 15 04:06:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	394515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	720838	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	667547	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	245768	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	269520	50.960	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.920%
35) Dibromofluoromethane	8.177	113	274028	56.001	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.000%
50) Toluene-d8	10.577	98	819820	44.350	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.700%
62) 4-Bromofluorobenzene	12.853	95	290932	51.096	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.200%
Target Compounds						
						Qvalue
3) Chloromethane	2.377	50	1941	0.369	ug/l #	79
16) Acetone	4.442	43	2274338	1380.485	ug/l	98
25) 2-Butanone	7.506	43	88196	34.008	ug/l #	87
52) Toluene	10.635	92	4413	0.340	ug/l	89
95) Naphthalene	15.641	128	11579	1.203	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071423\
Data File : VN078530.D
Acq On : 14 Jul 2023 18:27
Operator : JC\MD
Sample : 03637-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
A508

Quant Time: Jul 15 04:06:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

