

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071423\
 Data File : VN078514.D
 Acq On : 14 Jul 2023 11:56
 Operator : JC\MD
 Sample : 03624-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 15 03:59:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	399967	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	728085	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	686120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	221700	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	278847	52.004	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.000%
35) Dibromofluoromethane	8.177	113	277414	56.129	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.260%
50) Toluene-d8	10.576	98	861610	46.147	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.300%
62) 4-Bromofluorobenzene	12.853	95	294794	51.259	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.520%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071423\
Data File : VN078514.D
Acq On : 14 Jul 2023 11:56
Operator : JC\MD
Sample : 03624-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 15 03:59:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 1106
Delta R.T. 0.005 min
Lab File: VN078514.D
Acq: 14 Jul 2023 11:56




Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:168 Resp: 399967
Ion Ratio Lower Upper
168 100
99 59.1 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 52.004 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. 0.005 min
Lab File: VN078514.D
Acq: 14 Jul 2023 11:56

Tgt Ion: 65 Resp: 278847
Ion Ratio Lower Upper
65 100
67 51.9 0.0 106.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.112 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN078514.D

Acq: 14 Jul 2023 11:56

Instrument :

MSVOA_N

ClientSampleId :

STORAGE-BLANK-SOIL-REF-6





Tgt Ion:114 Resp: 728085

Ion Ratio Lower Upper

114 100

63 19.9 0.0 36.4

88 15.1 0.0 30.6



#35

Dibromofluoromethane

Concen: 56.129 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. -0.000 min

Lab File: VN078514.D

Acq: 14 Jul 2023 11:56




Tgt Ion:113 Resp: 277414


Ion Ratio Lower Upper

113 100

111 101.9 82.6 123.8

192 17.5 12.7 19.1









#50
Toluene-d8
Concen: 46.147 ug/l
RT: 10.576 min Scan# 1465
Delta R.T. 0.005 min
Lab File: VN078514.D
Acq: 14 Jul 2023 11:56




Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion: 98 Resp: 861610
Ion Ratio Lower Upper
98 100
100 62.3 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 51.259 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VN078514.D
Acq: 14 Jul 2023 11:56

Tgt Ion: 95 Resp: 294794
Ion Ratio Lower Upper
95 100
174 77.9 0.0 143.8
176 75.4 0.0 135.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.876 min Scan# 1

Delta R.T. 0.005 min

Lab File: VN078514.D

Acq: 14 Jul 2023 11:56

Instrument :

MSVOA_N

ClientSampleId :

STORAGE-BLANK-SOIL-REF-6





Tgt Ion:117 Resp: 686120

Ion Ratio Lower Upper

117 100

82 52.0 44.0 66.0

119 31.8 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.800 min Scan# 2014

Delta R.T. -0.000 min

Lab File: VN078514.D

Acq: 14 Jul 2023 11:56

Tgt Ion:152 Resp: 221700

Ion Ratio Lower Upper

152 100

115 62.1 32.3 96.8

150 161.5 0.0 355.8

