

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
 Data File : VN056704.D
 Acq On : 15 Jul 2019 12:10
 Operator : JC/SP
 Sample : VN0715WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0715WBL01

Quant Time: Jul 16 04:44:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 01:25:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	285953	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	499773	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	515804	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	192011	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	164447	51.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.98%	
35) Dibromofluoromethane	7.59	113	148580	47.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.54%	
50) Toluene-d8	10.09	98	626555	52.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.32%	
62) 4-Bromofluorobenzene	12.40	95	224731	55.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.96%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
Data File : VN056704.D
Acq On : 15 Jul 2019 12:10
Operator : JC/SP
Sample : VN0715WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0715WBL01

Quant Time: Jul 16 04:44:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 01:25:25 2019
Response via : Initial Calibration

