

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
 Data File : VN056714.D
 Acq On : 15 Jul 2019 16:02
 Operator : JC/SP
 Sample : PB121318TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121318TB

Quant Time: Jul 16 04:58:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 01:25:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	282387	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	487734	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	518808	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	196060	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	162173	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
35) Dibromofluoromethane	7.59	113	145601	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
50) Toluene-d8	10.09	98	625325	53.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.68%	
62) 4-Bromofluorobenzene	12.40	95	229009	57.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.86%	
Target Compounds						
16) Acetone	3.82	43	23347	16.220	ug/l	Qvalue 100
20) Methylene Chloride	4.56	84	6613	2.102	ug/l	94
43) Isopropyl Acetate	8.17	43	19787	2.830	ug/l #	90
95) Naphthalene	15.13	128	6789	5.113	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
Data File : VN056714.D
Acq On : 15 Jul 2019 16:02
Operator : JC/SP
Sample : PB121318TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB121318TB

Quant Time: Jul 16 04:58:21 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 01:25:25 2019
Response via : Initial Calibration

