

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
 Data File : VN056721.D
 Acq On : 15 Jul 2019 18:45
 Operator : JC/SP
 Sample : K3817-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-TW532-2126

Quant Time: Jul 16 09:10:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 01:25:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	269009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	475613	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	489600	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	191092	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	155748	52.34	ug/l	0.00
Spiked Amount						
			Recovery	=	104.68%	
35) Dibromofluoromethane	7.59	113	142266	48.07	ug/l	0.00
Spiked Amount						
			Recovery	=	96.14%	
50) Toluene-d8	10.09	98	586243	51.28	ug/l	0.00
Spiked Amount						
			Recovery	=	102.56%	
62) 4-Bromofluorobenzene	12.40	95	217479	56.41	ug/l	0.00
Spiked Amount						
			Recovery	=	112.82%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071519\
Data File : VN056721.D
Acq On : 15 Jul 2019 18:45
Operator : JC/SP
Sample : K3817-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-TW532-2126

Quant Time: Jul 16 09:10:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 01:25:25 2019
Response via : Initial Calibration

