

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
 Data File : VN056732.D
 Acq On : 15 Jul 2019 23:48
 Operator : JC/SP
 Sample : PB121354ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121354ZHE#02

Quant Time: Jul 16 05:38:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 01:25:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	269904	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	476222	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	511443	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	207707	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	157541	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
35) Dibromofluoromethane	7.59	113	142327	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
50) Toluene-d8	10.09	98	609925	53.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.58%	
62) 4-Bromofluorobenzene	12.40	95	227305	58.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.78%	
Target Compounds						
16) Acetone	3.82	43	8534	6.203	ug/l #	84
43) Isopropyl Acetate	8.17	43	17497	2.563	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071519\
Data File : VN056732.D
Acq On : 15 Jul 2019 23:48
Operator : JC/SP
Sample : PB121354ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB121354ZHE#02

Quant Time: Jul 16 05:38:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 01:25:25 2019
Response via : Initial Calibration

