

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
 Data File : VN056735.D
 Acq On : 16 Jul 2019 00:57
 Operator : JC/SP
 Sample : PB121354ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121354ZHE#05

Quant Time: Jul 16 05:46:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 01:25:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	274405	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	491476	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	526860	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	221631	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	163929	54.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.02%	
35) Dibromofluoromethane	7.59	113	147262	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
50) Toluene-d8	10.09	98	624758	52.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.78%	
62) 4-Bromofluorobenzene	12.40	95	240020	60.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.50%	
Target Compounds						
16) Acetone	3.83	43	8580	6.134	ug/l	95
43) Isopropyl Acetate	8.17	43	16930	2.403	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
Data File : VN056735.D
Acq On : 16 Jul 2019 00:57
Operator : JC/SP
Sample : PB121354ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB121354ZHE#05

Quant Time: Jul 16 05:46:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 01:25:25 2019
Response via : Initial Calibration

