

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
 Data File : VN056739.D
 Acq On : 16 Jul 2019 2:30
 Operator : JC/SP
 Sample : K3814-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KY0300F006-SW-190712

Quant Time: Jul 16 09:15:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 01:25:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	267400	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	475211	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	495010	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	194317	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	152648	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
35) Dibromofluoromethane	7.59	113	142764	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	10.09	98	589886	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
62) 4-Bromofluorobenzene	12.40	95	218172	56.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.28%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071519\
Data File : VN056739.D
Acq On : 16 Jul 2019 2:30
Operator : JC/SP
Sample : K3814-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KY0300F006-SW-190712

Quant Time: Jul 16 09:15:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 01:25:25 2019
Response via : Initial Calibration

