

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
 Data File : VN056711.D
 Acq On : 15 Jul 2019 14:53
 Operator : JC/SP
 Sample : K3797-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 16 04:53:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 01:25:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	279785	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	485062	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	500155	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	182267	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	160263	51.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.56%	
35) Dibromofluoromethane	7.59	113	146270	48.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.92%	
50) Toluene-d8	10.09	98	604276	51.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.66%	
62) 4-Bromofluorobenzene	12.40	95	217972	55.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.88%	


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071519\
Data File : VN056711.D
Acq On : 15 Jul 2019 14:53
Operator : JC/SP
Sample : K3797-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 16 04:53:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 01:25:25 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN056711.D

Acq: 15 Jul 2019 14:53

Instrument :

MSVOA_N

ClientSampleId :


STORAGE-BLANK-WATER-REF-4

Tot Ion:168 Resp: 279785

Ion Ratio Lower Upper

168 100

99 49.0 39.4 59.2



Abundance Ion 167.90 (167.60 to 168.60): VN056711.D (-1814) (-)

Ion 98.90 (98.60 to 99.60): VN056711.D (-1814) (-)

7.67



100000

50000

0

Time-->

7.60 7.70 7.80



#33

1,2-Dichloroethane-d4

Concen: 51.784 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. -0.00 min

Lab File: VN056711.D


Acq: 15 Jul 2019 14:53

Tgt Ion: 65 Resp: 160263

Ion Ratio Lower Upper

65 100

67 55.4 0.0 109.0



Abundance Ion 64.90 (64.60 to 65.60): VN056711.D (-1852) (-)

Ion 66.90 (66.60 to 67.60): VN056711.D (-1852) (-)

8.03

80000

60000


40000


20000

0

Time-->

7.95 8.00 8.05 8.10









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN056711.D
 Acq: 15 Jul 2019 14:53

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Tot Ion:114 Resp: 485062
 Ion Ratio Lower Upper
 114 100
 63 19.5 0.0 39.4
 88 15.4 0.0 31.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 48.456 ug/l
 RT: 7.59 min Scan# 1809
 Delta R.T. -0.00 min
 Lab File: VN056711.D
 Acq: 15 Jul 2019 14:53

Tgt Ion:113 Resp: 146270
 Ion Ratio Lower Upper
 113 100
 111 102.1 81.8 122.6
 192 22.3 17.5 26.3



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.830 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN056711.D

Acq: 15 Jul 2019 14:53

Instrument :

MSVOA_N

ClientSampleId :


STORAGE-BLANK-WATER-REF-4

Tot Ion: 98 Resp: 604276

Ion Ratio Lower Upper

98 100

100 64.5 52.0 78.0



Abundance Ion 98.00 (97.70 to 98.70): VN056700.D (-2570) (-)

Ion 100.00 (99.70 to 100.70): VN056700.D (-2570) (-)

10.09



300000

200000

100000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 55.440 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.00 min

Lab File: VN056711.D

Acq: 15 Jul 2019 14:53


Tgt Ion: 95 Resp: 217972

Ion Ratio Lower Upper

95 100

174 91.0 0.0 180.0

176 87.7 0.0 173.2



Abundance Ion 94.90 (94.60 to 95.60): VN056700.D (-3290) (-)

Ion 173.80 (173.50 to 174.50): VN056700.D (-3290) (-)

Ion 175.80 (175.50 to 176.50): VN056700.D (-3290) (-)

12.40


150000


100000

50000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN056711.D
Acq: 15 Jul 2019 14:53

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Tot Ion:117 Resp: 500155
Ion Ratio Lower Upper
117 100
82 53.3 42.2 63.2
119 32.6 25.2 37.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. -0.00 min
Lab File: VN056711.D
Acq: 15 Jul 2019 14:53

Tgt Ion:152 Resp: 182267
Ion Ratio Lower Upper
152 100
115 55.3 28.2 84.7
150 157.6 0.0 345.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

