

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071520\
 Data File : VN062528.D
 Acq On : 15 Jul 2020 11:31
 Operator : JC/MD
 Sample : VN0715WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0715WBL01

Quant Time: Jul 16 03:35:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	189578	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	333260	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	318634	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	117434	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	144143	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
35) Dibromofluoromethane	7.55	113	108256	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
50) Toluene-d8	10.06	98	433740	53.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.68%	
62) 4-Bromofluorobenzene	12.38	95	150337	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
Target Compounds						
20) Methylene Chloride	4.50	84	13735	5.189	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071520\
Data File : VN062528.D
Acq On : 15 Jul 2020 11:31
Operator : JC/MD
Sample : VN0715WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0715WBL01

Quant Time: Jul 16 03:35:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

