

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071520\
 Data File : VN062538.D
 Acq On : 15 Jul 2020 16:08
 Operator : JC/MD
 Sample : L3187-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-03-071420

Quant Time: Jul 16 02:31:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	190565	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	348935	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	341608	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	124252	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	146083	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
35) Dibromofluoromethane	7.55	113	109399	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
50) Toluene-d8	10.06	98	454039	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
62) 4-Bromofluorobenzene	12.38	95	162141	52.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.22%	
Target Compounds						
8) Diethyl Ether	3.37	74	30519	21.060	ug/l	95
16) Acetone	3.78	43	484336	389.203	ug/l	97
18) Methyl Acetate	4.29	43	4569	1.471	ug/l #	91
20) Methylene Chloride	4.50	84	1320642	496.350	ug/l	95
43) Isopropyl Acetate	8.13	43	74032	11.401	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071520\
Data File : VN062538.D
Acq On : 15 Jul 2020 16:08
Operator : JC/MD
Sample : L3187-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-03-071420

Quant Time: Jul 16 02:31:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

