

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071520\
 Data File : VN062539.D
 Acq On : 15 Jul 2020 16:32
 Operator : JC/MD
 Sample : L3188-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-071420

Quant Time: Jul 16 03:02:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	185823	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	338115	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	332330	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	121041	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	144123	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
35) Dibromofluoromethane	7.55	113	107397	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
50) Toluene-d8	10.06	98	443443	53.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.50%	
62) 4-Bromofluorobenzene	12.38	95	159236	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
Target Compounds						
8) Diethyl Ether	3.37	74	26837	18.992	ug/l	96
16) Acetone	3.78	43	398424	328.336	ug/l	97
18) Methyl Acetate	4.28	43	5376	1.775	ug/l #	88
20) Methylene Chloride	4.50	84	1172079	451.756	ug/l	94
43) Isopropyl Acetate	8.13	43	74330	11.813	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071520\
Data File : VN062539.D
Acq On : 15 Jul 2020 16:32
Operator : JC/MD
Sample : L3188-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-05-071420

Quant Time: Jul 16 03:02:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

