



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071521\
 Data File : VN067755.D
 Acq On : 15 Jul 2021 10:01
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N071421W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Wed Jul 14 01:19:19 2021



AutoFind: Averaged scan 3991 to 4030; Bkg corrected with scan 3990

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	5464	PASS
75	95	30	60	57.6	14123	PASS
95	95	100	100	100.0	24524	PASS
96	95	5	9	6.6	1617	PASS
173	174	0.00	2	0.1	13	PASS
174	95	50	100	72.5	17772	PASS
175	174	5	9	8.5	1513	PASS
176	174	95	101	95.9	17037	PASS
177	176	5	9	6.4	1093	PASS