

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071522\
 Data File : VN073475.D
 Acq On : 15 Jul 2022 10:40
 Operator : JC\MD
 Sample : VN0715MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0715MBL01

Quant Time: Jul 16 04:11:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	235081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	393218	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	361347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	163499	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	158778	46.787	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.580%
35) Dibromofluoromethane	7.951	113	133170	50.903	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.800%
50) Toluene-d8	10.375	98	424703	45.708	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.420%
62) 4-Bromofluorobenzene	12.669	95	170857	47.849	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071522\
Data File : VN073475.D
Acq On : 15 Jul 2022 10:40
Operator : JC\MD
Sample : VN0715MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0715MBL01

Quant Time: Jul 16 04:11:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

