

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071522\
 Data File : VN073476.D
 Acq On : 15 Jul 2022 11:04
 Operator : JC\MD
 Sample : VN0715WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0715WBL01

Quant Time: Jul 16 04:12:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	270133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	463037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	431249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	190027	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	197644	50.682	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.360%
35) Dibromofluoromethane	7.951	113	163329	53.018	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.040%
50) Toluene-d8	10.380	98	522465	47.751	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.500%
62) 4-Bromofluorobenzene	12.668	95	216402	51.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071522\
Data File : VN073476.D
Acq On : 15 Jul 2022 11:04
Operator : JC\MD
Sample : VN0715WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0715WBL01

Quant Time: Jul 16 04:12:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

