

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071522\
 Data File : VN073474.D
 Acq On : 15 Jul 2022 09:58
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 16 04:11:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.731	0.701	4.1	72	0.00
3 P	Chloromethane	0.710	0.696	2.0	78	0.00
4 C	Vinyl Chloride	0.603	0.582	3.5#	75	0.00
5 T	Bromomethane	0.285	0.384	-34.7#	109	-0.01
6 T	Chloroethane	0.355	0.366	-3.1	81	-0.02
7 T	Trichlorofluoromethane	1.043	1.056	-1.2	77	0.00
8 T	Diethyl Ether	0.424	0.449	-5.9	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.616	0.636	-3.2	82	0.00
10 T	Methyl Iodide	0.614	0.661	-7.7	72	0.00
11 T	Tert butyl alcohol	0.175	0.170	2.9	71	0.00
12 CM	1,1-Dichloroethene	0.589	0.585	0.7#	77	0.00
13 T	Acrolein	0.111	0.077	30.6#	53	0.00
14 T	Allyl chloride	1.240	1.320	-6.5	83	0.00
15 T	Acrylonitrile	0.450	0.538	-19.6	89	0.00
16 T	Acetone	0.405	0.471	-16.3	88	0.00
17 T	Carbon Disulfide	1.526	1.277	16.3	63	0.00
18 T	Methyl Acetate	1.135	1.431	-26.1#	100	0.00
19 T	Methyl tert-butyl Ether	2.145	2.346	-9.4	84	0.00
20 T	Methylene Chloride	0.712	0.738	-3.7	83	0.00
21 T	trans-1,2-Dichloroethene	0.615	0.614	0.2	78	0.00
22 T	Diisopropyl ether	2.314	2.698	-16.6	89	0.00
23 T	Vinyl Acetate	2.072	2.315	-11.7	83	0.00
24 P	1,1-Dichloroethane	1.197	1.337	-11.7	86	0.00
25 T	2-Butanone	0.645	0.771	-19.5	88	0.00
26 T	2,2-Dichloropropane	1.107	1.069	3.4	79	0.00
27 T	cis-1,2-Dichloroethene	0.745	0.770	-3.4	79	0.00
28 T	Bromochloromethane	0.501	0.597	-19.2	84	0.00
29 T	Tetrahydrofuran	0.442	0.518	-17.2	88	0.00
30 C	Chloroform	1.206	1.359	-12.7#	86	0.00
31 T	Cyclohexane	1.182	1.120	5.2	78	0.00
32 T	1,1,1-Trichloroethane	1.075	1.137	-5.8	80	0.00
33 S	1,2-Dichloroethane-d4	0.722	0.730	-1.1	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.333	0.344	-3.3	80	0.00
36 T	1,1-Dichloropropene	0.512	0.543	-6.1	80	0.00
37 T	Ethyl Acetate	0.773	0.906	-17.2	87	0.00
38 T	Carbon Tetrachloride	0.529	0.550	-4.0	78	0.00
39 T	Methylcyclohexane	0.637	0.654	-2.7	77	0.00
40 TM	Benzene	1.580	1.738	-10.0	82	0.00
41 T	Methacrylonitrile	0.375	0.449	-19.7	96	0.00
42 TM	1,2-Dichloroethane	0.573	0.641	-11.9	83	0.00
43 T	Isopropyl Acetate	1.112	1.313	-18.1	86	0.00
44 TM	Trichloroethene	0.402	0.408	-1.5	76	0.00
45 C	1,2-Dichloropropane	0.403	0.481	-19.4#	88	0.00
46 T	Dibromomethane	0.273	0.302	-10.6	83	0.00
47 T	Bromodichloromethane	0.529	0.625	-18.1	85	0.00
48 T	Methyl methacrylate	0.515	0.579	-12.4	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071522\
 Data File : VN073474.D
 Acq On : 15 Jul 2022 09:58
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 16 04:11:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.011	-10.0	81	0.00
50 S	Toluene-d8	1.181	1.189	-0.7	78	0.00
51 T	4-Methyl-2-Pentanone	0.705	0.894	-26.8#	91	0.00
52 CM	Toluene	0.976	1.056	-8.2#	81	0.00
53 T	t-1,3-Dichloropropene	0.583	0.665	-14.1	81	0.00
54 T	cis-1,3-Dichloropropene	0.629	0.715	-13.7	82	0.00
55 T	1,1,2-Trichloroethane	0.402	0.449	-11.7	85	0.00
56 T	Ethyl methacrylate	0.650	0.744	-14.5	81	0.00
57 T	1,3-Dichloropropane	0.689	0.790	-14.7	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.290	0.339	-16.9	80	0.00
59 T	2-Hexanone	0.537	0.677	-26.1#	89	0.00
60 T	Dibromochloromethane	0.383	0.462	-20.6	84	0.00
61 T	1,2-Dibromoethane	0.401	0.455	-13.5	81	0.00
62 S	4-Bromofluorobenzene	0.454	0.494	-8.8	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.408	0.401	1.7	77	0.00
65 PM	Chlorobenzene	1.131	1.175	-3.9	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.416	0.450	-8.2	83	0.00
67 C	Ethyl Benzene	2.053	2.210	-7.6#	82	0.00
68 T	m/p-Xylenes	0.775	0.829	-7.0	82	0.00
69 T	o-Xylene	0.791	0.826	-4.4	81	0.00
70 T	Styrene	1.221	1.375	-12.6	84	0.00
71 P	Bromoform	0.304	0.347	-14.1	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	4.318	4.519	-4.7	83	0.00
74 T	N-amyl acetate	2.162	2.280	-5.5	82	0.00
75 P	1,1,2,2-Tetrachloroethane	1.433	1.570	-9.6	88	0.00
76 T	1,2,3-Trichloropropane	1.307	1.358	-3.9	90	0.00
77 T	Bromobenzene	1.004	1.006	-0.2	80	0.00
78 T	n-propylbenzene	4.797	5.339	-11.3	84	0.00
79 T	2-Chlorotoluene	3.036	3.180	-4.7	83	0.00
80 T	1,3,5-Trimethylbenzene	3.519	3.761	-6.9	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.477	0.514	-7.8	80	0.00
82 T	4-Chlorotoluene	2.939	3.151	-7.2	84	0.00
83 T	tert-Butylbenzene	3.155	3.291	-4.3	81	0.00
84 T	1,2,4-Trimethylbenzene	3.514	3.783	-7.7	84	0.00
85 T	sec-Butylbenzene	4.237	4.761	-12.4	85	0.00
86 T	p-Isopropyltoluene	3.422	3.860	-12.8	84	0.00
87 T	1,3-Dichlorobenzene	1.801	1.874	-4.1	82	0.00
88 T	1,4-Dichlorobenzene	1.804	1.868	-3.5	82	0.00
89 T	n-Butylbenzene	2.852	3.343	-17.2	87	0.00
90 T	Hexachloroethane	0.638	0.674	-5.6	79	0.00
91 T	1,2-Dichlorobenzene	1.835	1.924	-4.9	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.348	0.362	-4.0	76	0.00
93 T	1,2,4-Trichlorobenzene	0.955	1.014	-6.2	80	0.00
94 T	Hexachlorobutadiene	0.425	0.447	-5.2	83	0.00
95 T	Naphthalene	3.854	3.814	1.0	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071522\
Data File : VN073474.D
Acq On : 15 Jul 2022 09:58
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jul 16 04:11:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.985	1.058	-7.4	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6