

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071524\
 Data File : VN082923.D
 Acq On : 15 Jul 2024 17:28
 Operator : JC\MD
 Sample : P3250-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Jul 16 01:25:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	76510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	148735	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	152896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	57963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	58630	54.666	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.340%	
35) Dibromofluoromethane	8.177	113	48403	50.716	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.440%	
50) Toluene-d8	10.571	98	170148	48.268	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.540%	
62) 4-Bromofluorobenzene	12.853	95	54110	42.733	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 85.460%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071524\
Data File : VN082923.D
Acq On : 15 Jul 2024 17:28
Operator : JC\MD
Sample : P3250-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Jul 16 01:25:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

