

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071619\
 Data File : VN056778.D
 Acq On : 16 Jul 2019 18:52
 Operator : JC/SP
 Sample : K3623-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-02-071519-C

Quant Time: Jul 17 01:11:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	267315	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	479548	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	524542	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	198299	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	160791	54.38	ug/l	0.00
Spiked Amount			Recovery	=	108.76%	
35) Dibromofluoromethane	7.59	113	143215	47.99	ug/l	0.00
Spiked Amount			Recovery	=	95.98%	
50) Toluene-d8	10.09	98	626134	54.32	ug/l	0.00
Spiked Amount			Recovery	=	108.64%	
62) 4-Bromofluorobenzene	12.40	95	229171	58.96	ug/l	0.00
Spiked Amount			Recovery	=	117.92%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
14) Allyl chloride	4.55	41	1223	0.281	ug/l #	29
16) Acetone	3.82	43	14861	10.906	ug/l	99
18) Methyl Acetate	4.33	43	8075	1.698	ug/l #	73
20) Methylene Chloride	4.55	84	40121	13.474	ug/l	96
25) 2-Butanone	6.85	43	2905	1.458	ug/l	95
30) Chloroform	7.37	83	1167	0.234	ug/l	91
31) Cyclohexane	7.67	56	4202	0.889	ug/l #	1
36) 1,1-Dichloropropene	7.66	75	17786	4.008	ug/l #	43
37) Ethyl Acetate	6.85	43	2905	0.664	ug/l #	68
38) Carbon Tetrachloride	7.67	117	23008	5.529	ug/l #	16
43) Isopropyl Acetate	8.17	43	17220	2.504	ug/l #	90
45) 1,2-Dichloropropane	9.27	63	1676	0.464	ug/l #	45
48) Methyl methacrylate	9.18	41	8609	2.589	ug/l #	13
51) 4-Methyl-2-Pentanone	9.99	43	4049	0.909	ug/l	93
54) cis-1,3-Dichloropropene	9.90	75	44786	8.695	ug/l #	61
57) 1,3-Dichloropropane	10.76	76	10141	1.833	ug/l #	42
59) 2-Hexanone	10.76	43	139795	43.749	ug/l #	51
74) N-amyl acetate	11.88	43	12812	2.152	ug/l #	50
76) 1,2,3-Trichloropropane	12.40	75	106181	24.482	ug/l #	100
79) 2-Chlorotoluene	12.72	91	92	Below Cal	#	39
80) 1,3,5-Trimethylbenzene	12.73	105	2783	0.205	ug/l	95
81) trans-1,4-Dichloro-2-buten	12.40	75	106181	75.609	ug/l #	13
83) tert-Butylbenzene	13.04	119	172	Below Cal	#	27
90) Hexachloroethane	13.99	117	882	0.400	ug/l #	3
95) Naphthalene	15.13	128	2900	4.774	ug/l #	87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071619\
Data File : VN056778.D
Acq On : 16 Jul 2019 18:52
Operator : JC/SP
Sample : K3623-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-02-071519-C

Quant Time: Jul 17 01:11:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52

Instrument :

MSVOA_N

ClientSampleId :


EO-02-071519-C

Tot Ion:168 Resp: 267315

Ion Ratio Lower Upper

168 100

99 50.6 39.4 59.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67

120000

100000

80000



60000

40000

20000

0

Time-->



#14

Allyl chloride

Concen: 0.281 ug/l

RT: 4.55 min Scan# 864

Delta R.T. 0.23 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52


Tgt Ion: 41 Resp: 1223

Ion Ratio Lower Upper

41 100

39 0.0 46.9 70.3#

76 0.0 26.5 39.7#



Abundance Ion 41.00 (40.70 to 41.70): VN

Ion 39.00 (38.70 to 39.70): VN

Ion 75.90 (75.60 to 76.60): VN


600


400

200

0

Time-->





#16

Acetone

Concen: 10.906 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.01 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52

Instrument :

MSVOA_N

ClientSampleId :





EO-02-071519-C

Tot Ion: 43 Resp: 14861

Ion Ratio Lower Upper

43 100

58 34.1 26.7 40.1



#18

Methyl Acetate

Concen: 1.698 ug/l

RT: 4.33 min Scan# 796

Delta R.T. 0.01 min

Lab File: VN056778.D




Acq: 16 Jul 2019 18:52


Tgt Ion: 43 Resp: 8075

Ion Ratio Lower Upper

43 100

74 10.6 18.9 28.3#









#20
Methvlene Chloride
Concen: 13.474 ug/l
RT: 4.55 min Scan# 865
Delta R.T. 0.01 min
Lab File: VN056778.D
Acq: 16 Jul 2019 18:52




Instrument :
MSVOA_N
ClientSampleId :
EO-02-071519-C


Tot Ion:	84	Resp:	40121
Ion	Ratio	Lower	Upper
84	100		
49	117.4	98.2	147.4
51	39.4	29.6	44.4
86	65.3	51.0	76.4



#25
2-Butanone
Concen: 1.458 ug/l
RT: 6.85 min Scan# 1580
Delta R.T. 0.02 min
Lab File: VN056778.D
Acq: 16 Jul 2019 18:52

Tgt Ion:	43	Resp:	2905
Ion	Ratio	Lower	Upper
43	100		
72	28.9	21.2	31.8









#30
Chloroform
Concen: 0.234 ug/l
RT: 7.37 min Scan# 1742
Delta R.T. 0.00 min
Lab File: VN056778.D
Acq: 16 Jul 2019 18:52




Instrument :
MSVOA_N
ClientSampleId :
EO-02-071519-C


Tot Ion: 83 Resp: 1167
Ion Ratio Lower Upper
83 100
85 57.8 51.9 77.9



#31
Cyclohexane
Concen: 0.889 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. 0.01 min
Lab File: VN056778.D
Acq: 16 Jul 2019 18:52

Tgt Ion: 56 Resp: 4202
Ion Ratio Lower Upper
56 100
69 172.4 25.8 38.8#
84 167.0 67.4 101.0#









#33
 1,2-Dichloroethane-d4
 Concen: 54.379 ug/l
 RT: 8.03 min Scan# 1945
 Delta R.T. -0.00 min
 Lab File: VN056778.D
 Acq: 16 Jul 2019 18:52




Instrument :
 MSVOA_N
 ClientSampleId :
 EO-02-071519-C


Tot Ion: 65 Resp: 160791
 Ion Ratio Lower Upper
 65 100
 67 55.8 0.0 109.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN056778.D
 Acq: 16 Jul 2019 18:52

Tgt Ion: 114 Resp: 479548
 Ion Ratio Lower Upper
 114 100
 63 19.7 0.0 39.4
 88 15.3 0.0 31.2






#35
 Dibromofluoromethane
 Concen: 47.989 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN056778.D
 Acq: 16 Jul 2019 18:52

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-02-071519-C

Tot Ion	Ratio	Lower	Upper
113	100		
111	104.3	81.8	122.6
192	23.1	17.5	26.3






Abundance

Ion 112.80 (112.50 to 113.50): V


Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36
 1,1-Dichloropropene
 Concen: 4.008 ug/l
 RT: 7.66 min Scan# 1832
 Delta R.T. -0.13 min
 Lab File: VN056778.D
 Acq: 16 Jul 2019 18:52

Tgt Ion	Ratio	Lower	Upper
75	100		
110	2.4	18.4	55.4#
77	0.0	24.6	36.8#






Abundance

Ion 74.90 (74.60 to 75.60): V

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): V









#37
Ethyl Acetate
Concen: 0.664 ug/l
RT: 6.85 min Scan# 1580
Delta R.T. -0.08 min
Lab File: VN056778.D
Acq: 16 Jul 2019 18:52

Instrument :
MSVOA_N
ClientSampled :
EO-02-071519-C

Tot Ion: 43 Resp: 2905
Ion Ratio Lower Upper
43 100
61 0.0 11.3 16.9#
70 0.0 8.5 12.7#




Abundance Ion 43.00 (42.70 to 43.70): VN0
Ion 60.90 (60.60 to 61.60): VN0
Ion 70.00 (69.70 to 70.70): VN0






#38
Carbon Tetrachloride
Concen: 5.529 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. -0.11 min
Lab File: VN056778.D
Acq: 16 Jul 2019 18:52

Tgt Ion: 117 Resp: 23008
Ion Ratio Lower Upper
117 100
119 5.3 77.5 116.3#
121 0.0 24.7 37.1#



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V





#43

Isopropyl Acetate

Concen: 2.504 ug/l

RT: 8.17 min Scan# 1988

Delta R.T. 0.00 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52

Instrument :


MSVOA_N

ClientSampleId :

EO-02-071519-C

Tot Ion: 43 Resp: 17220




Ion	Ratio	Lower	Upper
43	100		
61	20.6	22.3	33.5#
87	14.5	11.1	16.7



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#45

1,2-Dichloropropane

Concen: 0.464 ug/l

RT: 9.27 min Scan# 2332


Delta R.T. 0.15 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52



Tgt Ion: 63 Resp: 1676


Ion	Ratio	Lower	Upper
63	100		
65	0.0	24.0	36.0#



Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 64.90 (64.60 to 65.60): VN0









#48
Methvl methacrylate
Concen: 2.589 ug/l
RT: 9.18 min Scan# 2304
Delta R.T. -0.02 min
Lab File: VN056778.D
Acq: 16 Jul 2019 18:52

Instrument :
MSVOA_N
ClientSampled :
EO-02-071519-C

Tot Ion: 41 Resp: 8609
Ion Ratio Lower Upper
41 100
69 0.0 71.4 107.0#
39 0.0 41.2 61.8#




Abundance Ion 41.00 (40.70 to 41.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 39.00 (38.70 to 39.70): VN0






#50
Toluene-d8
Concen: 54.323 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN056778.D
Acq: 16 Jul 2019 18:52

Tgt Ion: 98 Resp: 626134
Ion Ratio Lower Upper
98 100
100 64.0 52.0 78.0



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0





#51

4-Methyl-2-Pentanone

Concen: 0.909 ug/l

RT: 9.99 min Scan# 2554

Delta R.T. 0.00 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52

Instrument :

MSVOA_N

ClientSampleId :


EO-02-071519-C

Tot Ion: 43 Resp: 4049

Ion Ratio Lower Upper




43 100

58 36.0 32.0 48.0



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#54

cis-1,3-Dichloropropene

Concen: 8.695 ug/l

RT: 9.90 min Scan# 2529

Delta R.T. 0.07 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52


Tgt Ion: 75 Resp: 44786

Ion Ratio Lower Upper

75 100

77 0.0 26.0 39.0#



39 28.3 35.1 52.7#




Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0

Ion 39.00 (38.70 to 39.70): VN0





#57

1,3-Dichloropropane

Concen: 1.833 ug/l

RT: 10.76 min Scan# 2795

Delta R.T. 0.05 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52

Instrument :

MSVOA_N

ClientSampleId :


EO-02-071519-C

Tot Ion: 76 Resp: 10141

Ion Ratio Lower Upper




76 100

78 0.0 26.0 39.0#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0



#59

2-Hexanone

Concen: 43.749 ug/l

RT: 10.76 min Scan# 2796

Delta R.T. 0.01 min

Lab File: VN056778.D


Acq: 16 Jul 2019 18:52

Tgt Ion: 43 Resp: 139795

Ion Ratio Lower Upper



43 100


58 19.8 27.9 83.6#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0





#62

4-Bromofluorobenzene

Concen: 58.958 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.00 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52

Instrument :

MSVOA_N

ClientSampled :

EO-02-071519-C


Tot Ion: 95 Resp: 229171

Ion Ratio Lower Upper

95 100

174 90.8 0.0 180.0




176 87.8 0.0 173.2



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): VN0

Ion 175.80 (175.50 to 176.50): VN0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. -0.00 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52


Tgt Ion: 117 Resp: 524542

Ion Ratio Lower Upper

117 100

82 53.0 42.2 63.2



119 32.3 25.2 37.8




Abundance Ion 116.90 (116.60 to 117.60): VN0

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): VN0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3598

Delta R.T. -0.00 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52

Instrument :

MSVOA_N

ClientSampleId :

EO-02-071519-C


Tot Ion:152 Resp: 198299

Ion Ratio Lower Upper

152 100

115 56.6 28.2 84.7




150 157.8 0.0 345.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 2.152 ug/l

RT: 11.88 min Scan# 3142

Delta R.T. -0.19 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52

Tgt Ion: 43 Resp: 12812


Ion Ratio Lower Upper

43 100

70 1.5 34.5 51.7#

55 3.0 20.2 30.2#

61 9.7 19.9 29.9#






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V









#76
 1,2,3-Trichloropropane
 Concen: 24.482 ug/l
 RT: 12.40 min Scan# 3305
 Delta R.T. -0.15 min
 Lab File: VN056778.D
 Acq: 16 Jul 2019 18:52




Instrument :
 MSVOA_N
 ClientSampleId :
 EO-02-071519-C


Tot Ion: 75 Resp: 106181
 Ion Ratio Lower Upper
 75 100
 77 1.1 0.0 0.0#



#79
 2-Chlorotoluene
 Concen: Below Cal
 RT: 12.72 min Scan# 3406
 Delta R.T. 0.05 min
 Lab File: VN056778.D
 Acq: 16 Jul 2019 18:52

Tgt Ion: 91 Resp: 92
 Ion Ratio Lower Upper
 91 100
 126 0.0 17.9 53.7#





#80

1,3,5-Trimethylbenzene

Concen: 0.205 ug/l

RT: 12.73 min Scan# 3408

Delta R.T. -0.00 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52

Instrument :

MSVOA_N

ClientSampleId :





EO-02-071519-C

Tot Ion:105 Resp: 2783

Ion Ratio Lower Upper

105 100

120 46.1 24.8 74.3



#81

trans-1,4-Dichloro-2-butene

Concen: 75.609 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.10 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52




Tgt Ion: 75 Resp: 106181


Ion Ratio Lower Upper

75 100

53 0.0 73.7 110.5#

89 0.0 36.2 54.2#









#83
 tert-Butylbenzene
 Concen: Below Cal
 RT: 13.04 min Scan# 3504
 Delta R.T. 0.05 min
 Lab File: VN056778.D
 Acq: 16 Jul 2019 18:52




Instrument :
 MSVOA_N
 ClientSampleId :
 EO-02-071519-C


Tot Ion:119 Resp: 172
 Ion Ratio Lower Upper
 119 100
 91 0.0 31.8 95.4#
 134 0.0 13.0 38.9#



#90
 Hexachloroethane
 Concen: 0.400 ug/l
 RT: 13.99 min Scan# 3801
 Delta R.T. 0.12 min
 Lab File: VN056778.D
 Acq: 16 Jul 2019 18:52

Tgt Ion:117 Resp: 882
 Ion Ratio Lower Upper
 117 100
 201 0.0 46.5 139.5#





#95

Naphthalene

Concen: 4.774 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN056778.D

Acq: 16 Jul 2019 18:52

Instrument :

MSVOA_N

ClientSampleId :

EO-02-071519-C



Tot Ion:128 Resp: 2900

Ion Ratio Lower Upper

128 100

127 10.0 10.3 15.5#

129 3.7 8.7 13.1#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

