

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071619\
 Data File : VN056761.D
 Acq On : 16 Jul 2019 12:17
 Operator : JC/SP
 Sample : K3836-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JOYCE-SW-7-12

Quant Time: Jul 17 08:20:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	261820	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	460422	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	477555	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	170714	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	152845	52.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.56%	
35) Dibromofluoromethane	7.59	113	139704	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
50) Toluene-d8	10.09	98	577206	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
62) 4-Bromofluorobenzene	12.40	95	206265	55.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.54%	
Target Compounds						
16) Acetone	3.82	43	26330	19.729	ug/l	96
51) 4-Methyl-2-Pentanone	9.99	43	13776	3.222	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071619\
Data File : VN056761.D
Acq On : 16 Jul 2019 12:17
Operator : JC/SP
Sample : K3836-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JOYCE-SW-7-12

Quant Time: Jul 17 08:20:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

