

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071619\
 Data File : VN056762.D
 Acq On : 16 Jul 2019 12:40
 Operator : JC/SP
 Sample : K3827-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 UTILITY-ROOM-BATHROOM

Quant Time: Jul 17 08:21:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	271400	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	481908	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	501243	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	175538	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	158053	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
35) Dibromofluoromethane	7.59	113	142705	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
50) Toluene-d8	10.09	98	607801	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
62) 4-Bromofluorobenzene	12.40	95	212665	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
Target Compounds						
29) Tetrahydrofuran	7.22	42	3609	2.736	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071619\
Data File : VN056762.D
Acq On : 16 Jul 2019 12:40
Operator : JC/SP
Sample : K3827-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
UTILITY-ROOM-BATHROOM

Quant Time: Jul 17 08:21:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

