


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN071619\
Data File : VN056752.D
Acq On : 16 Jul 2019 8:21
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Title : SW846 8260
Last Update : Tue Jul 16 07:03:34 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	9328	PASS
75	95	30	60	47.8	27056	PASS
95	95	100	100	100.0	56600	PASS
96	95	5	9	6.3	3564	PASS
173	174	0.00	2	0.8	426	PASS
174	95	50	100	89.8	50808	PASS
175	174	5	9	7.9	4039	PASS
176	174	95	101	97.9	49720	PASS
177	176	5	9	6.3	3139	PASS