

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071620\
 Data File : VN062546.D
 Acq On : 16 Jul 2020 14:34
 Operator : JC/MD
 Sample : L3321-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 200715087-01-TRIP-BLANK

Quant Time: Jul 16 18:26:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N071320W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Jul 14 00:52:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	36591	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	187741	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	167792	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	7.99	65	94554	30.08	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.27%
60) 4-Bromofluorobenzene	12.38	95	71389	25.00	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	83.33%
63) Toluene-d8	10.06	98	238427	30.07	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.23%

Target Compounds					Ovalue
15) Acetone	3.78	58	6717	16.649	ug/l 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071620\
Data File : VN062546.D
Acq On : 16 Jul 2020 14:34
Operator : JC/MD
Sample : L3321-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
200715087-01-TRIP-BLANK

Quant Time: Jul 16 18:26:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N071320W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Jul 14 00:52:57 2020
Response via : Initial Calibration

