

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071620\
 Data File : VN062550.D
 Acq On : 16 Jul 2020 16:10
 Operator : JC/MD
 Sample : L3320-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 200715090-01-TRIP-BLANK

Quant Time: Jul 16 18:39:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N071320W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Jul 14 00:52:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	37891	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	199673	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	179454	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	7.99	65	100224	30.79	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.63%
60) 4-Bromofluorobenzene	12.38	95	75264	24.65	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	82.17%
63) Toluene-d8	10.06	98	251957	29.71	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.03%

Target Compounds					Ovalue
15) Acetone	3.78	58	8982	21.499 ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071620\
Data File : VN062550.D
Acq On : 16 Jul 2020 16:10
Operator : JC/MD
Sample : L3320-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
200715090-01-TRIP-BLANK

Quant Time: Jul 16 18:39:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N071320W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Jul 14 00:52:57 2020
Response via : Initial Calibration

