

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071621\
 Data File : VN067775.D
 Acq On : 16 Jul 2021 16:09
 Operator : JC/MD
 Sample : M3051-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41213

Quant Time: Jul 17 04:58:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	348799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	647917	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	623545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	226873	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	332444	63.065	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 126.140%			
35) Dibromofluoromethane	8.021	113	238131	60.832	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 121.660%			
50) Toluene-d8	10.443	98	994546	60.697	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 121.400%			
62) 4-Bromofluorobenzene	12.733	95	326888	56.475	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 112.940%			
Target Compounds						
16) Acetone	4.293	43	6919	4.171	ug/l	92
52) Toluene	10.507	92	13142	1.114	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071621\
Data File : VN067775.D
Acq On : 16 Jul 2021 16:09
Operator : JC/MD
Sample : M3051-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41213

Quant Time: Jul 17 04:58:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

